



King's Research Portal

DOI:

[10.1016/j.pnmrs.2021.03.002](https://doi.org/10.1016/j.pnmrs.2021.03.002)

Document Version

Publisher's PDF, also known as Version of record

[Link to publication record in King's Research Portal](#)

Citation for published version (APA):

Ljungberg, E., Damestani, N. L., Wood, T. C., Lythgoe, D. J., Zelaya, F., Williams, S. C. R., Solana, A. B., Barker, G. J., & Wiesinger, F. (2021). Silent zero TE MR neuroimaging: Current state-of-the-art and future directions. *PROGRESS IN NUCLEAR MAGNETIC RESONANCE SPECTROSCOPY*, 123, 73-93.
<https://doi.org/10.1016/j.pnmrs.2021.03.002>

Citing this paper

Please note that where the full-text provided on King's Research Portal is the Author Accepted Manuscript or Post-Print version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version for pagination, volume/issue, and date of publication details. And where the final published version is provided on the Research Portal, if citing you are again advised to check the publisher's website for any subsequent corrections.


General rights

Copyright and moral rights for the publications made accessible in the Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognize and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the Research Portal

Take down policy

If you believe that this document breaches copyright please contact librarypure@kcl.ac.uk providing details, and we will remove access to the work immediately and investigate your claim.



Silent zero TE MR neuroimaging: Current state-of-the-art and future directions



Emil Ljungberg^{a,*}, Nikou L. Damestani^a, Tobias C. Wood^a, David J. Lythgoe^a, Fernando Zelaya^a, Steven C.R. Williams^a, Ana Beatriz Solana^b, Gareth J. Barker^a, Florian Wiesinger^{a,b}

^a Department of Neuroimaging, Institute of Psychiatry, Psychology & Neuroscience, King's College London, London, United Kingdom

^b ASL Europe, GE Healthcare, Munich, Germany

Edited by David Gadian and David Neuhaus

ARTICLE INFO

Article history:

Received 25 November 2020

Accepted 9 March 2021

Available online 26 March 2021

Keywords:

Zero Echo Time (ZTE)

Silent MRI

Neuroimaging

ABSTRACT

Magnetic Resonance Imaging (MRI) scanners produce loud acoustic noise originating from vibrational Lorentz forces induced by rapidly changing currents in the magnetic field gradient coils. Using zero echo time (ZTE) MRI pulse sequences, gradient switching can be reduced to a minimum, which enables near silent operation. Besides silent MRI, ZTE offers further interesting characteristics, including a nominal echo time of $TE = 0$ (thus capturing short-lived signals from MR tissues which are otherwise MR-invisible), 3D radial sampling (providing motion robustness), and ultra-short repetition times (providing fast and efficient scanning). In this work we describe the main concepts behind ZTE imaging with a focus on conceptual understanding of the imaging sequences, relevant acquisition parameters, commonly observed image artefacts, and image contrasts. We will further describe a range of methods for anatomical and functional neuroimaging, together with recommendations for successful implementation.

© 2021 The Author(s). Published by Elsevier B.V. This is an open access article under the CC BY license (<http://creativecommons.org/licenses/by/4.0/>).

Contents

1. Introduction	74
2. MRI with ZTE	75
2.1. RF excitation in the presence of gradients	75
2.2. Silent 3D radial k-space sampling	76
2.3. Native ZTE image contrast	76
2.4. Image artefacts	77
3. ZTE pulse sequences and control of image contrast	77
3.1. Magnetization prepared ZTE	77
3.2. Multi-echo gradient refocused ZTE	79
3.2.1. Looping Star	79
3.2.2. ZTE-BURST	80
4. ZTE for structural neuroimaging	80
4.1. T_1 contrast mechanisms	80
4.2. T_2 contrast mechanisms	81
4.3. Multiparametric ZTE	82
4.4. Diffusion contrast	82
4.5. Magnetization transfer	82
4.6. Angiography	83
4.7. Quantitative susceptibility mapping and T_2^*	85

* Corresponding author at: Centre for Neuroimaging Sciences, De Crespigny Park, London, SE5 8AF, United Kingdom.

E-mail address: emil.ljungberg@kcl.ac.uk (E. Ljungberg).

4.8. Ultra-short T_2 : Bone and myelin imaging	85
4.9. Other ZTE applications	87
5. ZTE for functional neuroimaging	87
6. Discussion	88
7. Conclusions	89
Declaration of Competing Interest	89
Acknowledgements	89
References	89

1. Introduction

Magnetic Resonance Imaging (MRI) scanners produce loud acoustic noise because of Lorentz forces caused by rapidly changing currents in the magnetic field gradient coils used primarily for spatial localization [1]. The acoustic noise produced with conventional sequences is typically around 90–110 dBA (where dBA is dB on the A-weighted scale, which accounts for the sensitivity of the human ear at different frequencies) [2–4]. For fast imaging sequences, such as Echo Planar Imaging (EPI), it can even reach levels up to 130 dBA [5]. Since the Lorentz force scales with the main magnetic field strength, acoustic noise also increases at higher field strength [6].

It is generally acknowledged that the acoustic noise produced by the MRI scanner is one of the most unpleasant aspects of the scan experience for patients [7–12]. Given appropriate and correctly worn hearing protection [13,14] there are, to our knowledge, no studies showing permanent hearing loss after a single MRI scan [15,16]. However, studies have reported temporary effects on hearing following scans at both 1.5T [17] and 3T [3].

Exposure to the loud noise inside the MRI scanner for long periods of time is problematic for certain groups of individuals. For instance, to avoid motion artefacts in neonatal and paediatric MRI, scanning is preferably performed under natural sleep [18,19], and this requires reduction of the acoustic noise. For individuals with hyperacusis [20], i.e., perception of ordinary sounds as abnormally loud, the noise from the MRI scanning can cause discomfort. Studies have found hyperacusis to be prevalent in

numerous conditions including tinnitus, migraines, and autism spectrum disorder [20–22].

In functional MRI (fMRI) studies, the acoustic noise is an additional confounding sensory stimulus, and can impact the blood-oxygen level dependent (BOLD) response as a function of both its loudness [23] and duration [24]. The effect of acoustic noise on the BOLD activation appears to vary based on the task performed [25–27], and resting state network identification can be impacted by the sparse-sampling technique employed [28,29]. There are also patient groups with auditory symptoms who are difficult to study using conventional fMRI, since the information related to the auditory stimulus of the study has to be extracted from the background noise, as has been described in studies of tinnitus [30] and Williams syndrome [31].

The acoustic noise from the MRI scanner can be reduced through hardware modifications, such as novel gradient designs and shielding [32–34]. Most MRI manufacturers also support a “quiet scanning” mode, typically using standard pulse sequences but with derated gradient performance for smoother temporal changes [35,36]. Despite all these improvements, the vast majority of MRI scans performed today still remain very loud and require the patient to wear hearing protection for additional noise suppression, with typical earplugs reducing the noise by 20–30 dBA [13].

Instead of lowering the acoustic noise of loud MRI sequences, it is also possible to diminish the generation of acoustic noise in the first place by minimising gradient switching, which can be achieved with Zero Echo Time (ZTE) pulse sequences [37,38]. The incremental development leading to silent ZTE imaging is best




Fig. 1. Simplified pulse sequences diagrams showing RF excitation and gradients in one dimension for (A) UTE, (B) BLAST and (C) RUFIS. In UTE imaging, RF excitation is performed prior to the readout gradients. In the BLAST pulse sequence, gradients are ramped up before RF excitation and ramped down after readout. In RUFIS, RF excitation is performed with the gradient on, as in BLAST, but without returning gradients to zero between excitations, hence minimizing gradient switching and allowing silent imaging.

appreciated through comparison to the ultra-short echo time (UTE) pulse sequence [39,40], shown in Fig. 1. Both UTE and ZTE sequences are variants of a pulse-acquire free induction decay (FID) pulse sequence which do not create echo signals, and thus the term echo time (TE) can be confusing. When used for imaging, however, the delay between the middle of the RF pulse and the time at which the central k-space point (k_0) is acquired determines the degree of T_2^* weighting, directly analogous to the TE in a gradient echo, and therefore the term TE will still be used.

In a UTE pulse sequence as shown in Fig. 1A, RF excitation is performed before gradients are ramped up and data acquisition begins after a short, non-negligible, delay time. Thus, k_0 is acquired at a non-zero TE. The rapid gradient switching and simultaneous readout make UTE acoustically loud and susceptible to gradient delay and eddy current artefacts [41–43].

In ZTE imaging, first demonstrated with the Back-projection Low Angle Shot (BLAST) sequence [44] shown in Fig. 1B, the RF pulse is shifted to be applied after the gradient has reached the target amplitude, which means that k_0 effectively is acquired at TE = 0. Hardware constraints, described in Section 2.1, makes it difficult to actually acquire this point, but the samples that are acquired are consistent with TE = 0 and thus the term ZTE is appropriate. By acquiring the FID during a constant gradient, ZTE imaging avoids the issues with the ramp-sampling characteristic in UTE imaging. However, similar to UTE imaging, BLAST will produce high acoustic noise due to the rapid gradient switching required for efficient imaging.

The ZTE Rotating Ultra-Fast Imaging Sequence (RUFIS) [45], shown in Fig. 1C, forms the basis for silent imaging with ZTE. It is similar to BLAST in that RF excitation is performed with the gradient on, but between excitations the readout gradient is ramped straight to the value required for the next set of sample points, for a faster acquisition and reduced gradient switching. The motivation of the original RUFIS method was ultra-fast imaging without extreme demands on the gradient system, but no specific emphasis was given to its potential for silent imaging.

An alternative to the hard pulse ZTE sequence shown in Fig. 1 is SWIFT (SWEEP Imaging with Fourier Transformation) which uses swept RF pulse excitation [46,47]. SWIFT has been used less frequently than hard pulse ZTE for human neuroimaging applications primarily because of very demanding RF transmit-receive switching requirements. It has also been shown by Weiger et al. that while sweep excitation can produce higher flip angles, the SNR efficiency is actually equivalent to hard pulse excitation [47].

The purpose of this review is to give a comprehensive description of silent neuroimaging using ZTE MR pulse sequences. In the first half, the concept of ZTE will be reviewed from an MR physics point of view, together with descriptions of variations and extensions of the native ZTE pulse sequence. In the second half, practical neuroimaging applications of ZTE will be reviewed. We conclude with an outlook on the future of silent ZTE and potential translation to the clinic.

2. MRI with ZTE

This section describes the mechanics of the ZTE pulse sequence in terms of RF excitation and the k-space encoding. It lists common ZTE image artefacts and describes the acoustic noise behaviour of ZTE.

2.1. RF excitation in the presence of gradients

As described in Fig. 1, ZTE imaging requires RF excitation in the presence of the readout gradients, which leads to two unique challenges. Firstly, the excitation bandwidth of the RF pulse has to

encompass the full receive imaging bandwidth to ensure uniform excitation independent of the readout direction [48]. Secondly, the finite time it takes to change from RF transmit to receive mode, typically referred to as the dead-time gap, causes some samples to be missed at the beginning of the readout [49].

A block RF pulse (i.e., rectangular) with pulse width τ_{TX} produces a sinc-shaped excitation profile along the readout direction according to

$$P(\vec{r}) = \alpha \cdot \text{sinc}(\gamma \cdot \tau_{TX} \cdot \vec{G} \cdot \vec{r}) \quad \text{with} \quad \text{sinc}(x) = \frac{\sin(\pi x)}{\pi x} \quad (1)$$

where $\alpha = \gamma B_1 \tau_{TX}$ is the nominal flip angle, γ is the gyromagnetic ratio, B_1 is the RF excitation amplitude, \vec{G} is the spoke gradient vector, and \vec{r} is the position vector [48]. The excitation profile rotates in synchrony with the readout spoke direction and the effective flip angle progressively deviates from the nominal flip angle (α) with increasing distance from the isocentre. In order to avoid the first zero crossing of the sinc-shaped excitation profile falling within the field of view (FOV) (i.e., $\gamma \cdot \tau_{TX} \cdot |\vec{G}| \cdot \text{FOV}/2 < 1$), the excitation pulse width needs to be shorter than twice the dwell time (dw)

$$\tau_{TX} < 2 \cdot dw, \quad \text{with} \quad dw = \frac{1}{BW_{RX}}, \quad \text{and} \quad BW_{RX} = \gamma \cdot |\vec{G}| \cdot \text{FOV} \quad (2)$$

with BW_{RX} being the imaging bandwidth. This condition limits the maximum flip angle (α_{max}) proportional to the maximum RF excitation field ($B_{1,max}$) and inverse proportional to the imaging bandwidth according to $\alpha_{max} = 2 \cdot \gamma \cdot B_{1,max}/BW_{RX}$. For example, assuming $B_{1,max} = 15 \mu\text{T}$ and $BW_{RX} = \pm 62.5 \text{ kHz}$, the maximum flip angle is limited to $\alpha_{max} \approx 3.7^\circ$.

For a sufficiently large number of uniformly distributed spokes the effective excitation profile can, to a good approximation, be assumed to be spherically symmetric. However, inconsistent excitation dependent on the readout direction leads to blurring which increases with distance from isocentre. This can be corrected to a limited extent by correcting each spoke by the effective excitation profile as described by Grodzki et al. [48]. Alternatively, shaped RF pulses can be used to improve excitation uniformity in the presence of the readout gradient, but at the expense of lower maximum flip angle for the same B_1 amplitude, as well as high demands on the time resolution of the RF transmit system [50]. It is also possible to reduce the excitation profile effects by performing RF excitation with reduced gradient amplitude and then increasing the amplitude for data acquisition, resulting in hybrid ramp sampling [51–53]. While this method enables excitation with higher flip angles, it will increase the acoustic noise due to increased gradient switching [51]. The acquisition will also be more sensitive to eddy currents from ramp-sampling, thus requiring gradient calibration for accurate image encoding [53].

When RF excitation is applied in the presence of the readout gradient, k-space encoding starts immediately (hence TE = 0). Given the finite and non-zero switching time between RF transmit and receive mode, there is always a small dead-time gap (Δt) as shown in Fig. 2A in the beginning of each spoke, resulting in a number of missed data points. This translates to a spherical region in the centre of k-space without acquired k-space samples. The number of radial samples missed (n_Δ) during the dead-time gap (Δt) can be estimated as

$$n_\Delta = \frac{\Delta t}{dw} \quad (3)$$

Several factors affect the dead-time gap Δt , including the RF excitation pulse width (τ_{TX}), as well as system-specific delays for the readout filter, and it is therefore difficult to estimate Δt without




Fig. 2. (A) Simplified ZTE pulse sequence diagram with two spokes, showing RF excitation and the gradient waveform on one axis together with magnification of the RF excitation part of the spoke, showing the dead-time gap Δt after RF excitation. (B) 3D view of spoke distribution in k-space with the endpoints of each spoke connected by the blue line.

detailed knowledge of the MR system [38]. As an example, with $\Delta t = 20 \mu\text{s}$ and $\text{BW}_{\text{RX}} = \pm 62.5 \text{ kHz}$, the dead-time gap results in $n_{\Delta} \approx 2.5$ missed centre k-space samples.

A limited number of missing samples can be recovered during reconstruction using linear algebra, provided that symmetric spokes have been acquired, and that the imaging object is of finite support (i.e., imaging FOV fully encompasses the MR active object) [49,54]. For larger dead-time gaps there are several methods that acquire additional data to replace the missing samples similar to keyhole imaging [55], including: WASPI (Water- And fat-Suppressed proton Projection MRI) [56], PETRA (Pointwise Encoding Time reduction with Radial Acquisition) [57] and HYFI (Hybrid Filling) [58]. WASPI acquires a second radial acquisition with reduced gradient strength, resulting in fewer missed samples, while PETRA acquires the missing samples pointwise on a Cartesian grid. HYFI combines both pointwise encoding and radial projections with different gradient strength. We note that the name WASPI actually derives from a specific application of the ZTE sequence described in the original paper by Wu et al. [56], and not the reduced gradient amplitude approach to dead-time gap filling with which it has now become synonymous.

2.2. Silent 3D radial k-space sampling

Sampling of the free induction decay (FID) signal in the presence of a constant readout gradient naturally leads to a 3D centre-out radial k-space sampling scheme. By updating the readout direction in between excitations, a full spherical k-space is sequentially encoded, as shown in Fig. 2B. For a cubic image matrix size of size $N \times N \times N$, each spoke contains $N_{\text{pts}} = N/2$ sampling points. The number of spokes required to fulfil the Nyquist criterion (N_{Nyquist}) at a maximum k-space radius is determined by the surface area of the k-space sphere, i.e., $N_{\text{Nyquist}} = \pi N^2$ [59]. For equidistant radial sampling, the density decreases inversely proportional to the squared radius. Accelerating a 3D radial acquisition through angular undersampling, i.e., reduction of the number of spokes, will reduce the SNR and produce undersampling artefacts manifesting as streaking.

The non-selective excitation in ZTE will excite spins outside the FOV, potentially including plastic materials such as the RF coil and patient table [37,60]. To avoid aliasing of such signals, radial oversampling is used to push the aliasing-sphere outside the imaging FOV [38]. Radial oversampling, resulting in a larger encoded FOV,

is also essential for algebraic reconstruction of the deadtime gap as it ensures that the object has finite support in image space [49].

The essential features which render the ZTE pulse sequence silent are 1) constant gradient FID readout and 2) minimal updates of the readout direction (i.e., not readout amplitude) in between repetitions. This does not mean that all ZTE acquisitions are necessarily silent, but rather that a ZTE pulse sequence can operate within these constraints. For instance, the ZTE BLAST sequence shown in Fig. 1B, is not silent due to large gradient steps between readouts. The small change in gradient direction between spokes in RUFIS, on the other hand, enables silent acquisition. For complete 3D k-space coverage this can be achieved by arranging the spokes in a spiral pattern, as shown in Fig. 2B [61]. The acoustic sound pressure and frequency spectrum produced by a ZTE sequence depend on scan parameters such as the readout bandwidth, TR and number of spokes in the trajectory, since these will affect absolute gradient amplitude, the duration of each spoke, and the gradient steps between spokes. With commonly used scan parameters for ZTE, the acoustic noise typically stays within 5 dB of ambient noise levels, as shown in Table 1 which summarizes acoustic noise measurements from published studies using ZTE for silent imaging.

There are several imaging pulse sequences in the literature that can be used for silent ZTE imaging, including RUFIS, PETRA, WASPI, and Looping Star. For simplicity, and to clarify that our descriptions generalise across these methods, for the remainder of this paper we use “ZTE” as an over-arching term for any FID pulse sequence with a $TE = 0$ FID readout, which also features small gradient steps between readouts to allow near-silent data acquisition.

2.3. Native ZTE image contrast

The contrast behaviour of an RF spoiled steady-state ZTE sequence is similar to that of a spoiled gradient echo (SPGR) acquisition (often also referred to as GRE, FLASH, or Fast Field-Echo) [75]. In an SPGR acquisition, the longitudinal steady-state magnetization can be described by the following equation

$$M_{z,\text{SPGR}} = PD \cdot E_2^* \cdot \frac{(1 - E_1)}{1 - E_1 \cdot \cos(\alpha)}$$

with $E_1 = e^{-\frac{TR}{T_1}}$, $E_2^* = e^{-\frac{TE}{T_2^*}}$ (4)

with the contrast depending on T_1 , T_2^* , the proton density (PD) and the flip angle α . When $TE = 0$, T_2^* decay is reduced to a minimum, i.e., $E_2^* = 1$, resulting in heightened sensitivity to tissues with

Table 1

Acoustic noise measurements with silent ZTE sequences from published studies with comparisons to conventional sequences, when available. Values have been rounded to the same precision for comparison, and differences between sequences and ambient noise levels were calculated before rounding.

Modality	B_0	L_{AMB}	Silent			Comparison				Unit	Ref
			Sequence	L_{ZTE}	$L_{ZTE}-L_{AMB}$	Sequence	L_{STD}	$L_{STD}-L_{AMB}$	$L_{STD}-L_{ZTE}$		
Structural T_1	3T	69	RUFIS	69	< 1	IR-SPGR	105	36	36	dB	[62]**
Structural T_1	3T	52	RUFIS	53	< 1	IR-SE	82	30	30	dB	[63]*
Structural T_1	7T	53	RUFIS	55	2	IR-SPGR	90	37	35	dBA	[64]*
Structural T_1	3T	48	PETRA	51	3	MPRAGE	78	30	27	dBA	[65]*
Structural T_1	3T	53	PETRA	58	5	MPRAGE	87	34	29	dBA	[66]*
VFA T_1	3T	70	RUFIS	70	< 1	SPGR	103	33	33	dBA	[67]**
MRA	3T	55	RUFIS	58	3	TOF	93	38	35	dB	[68]*
DWI	3T	51	DWI-RUFIS	54	3	DWI-EPI	85	34	31	dB	[69]*
T_2^* / QSM	3T	67	ZTE-BURST [†]	76	9	mGRE	103	37	27	dBA	[70]**
T_2^* / QSM	3T	64	Looping Star ^{††}	73	8	N/A	–	–	–	dBA	[71]**
fMRI	3T	64	Looping Star	67	3	N/A	–	–	–	dBA	[71]**
fMRI	3T	71	Looping Star	71	1	GRE-EPI	103	32	32	dBA	[72]**
fMRI	3T	N/A	Looping Star	74	–	GRE-EPI	108	–	34	dBA	[73]**
fMRI	3T	72	T_2 -prep RUFIS	75	2	GRE-EPI	114	42	39	dBA	[74]**

Abbreviations: B_0 : Main magnetic field strength, L_{AMB} : Ambient Sound Pressure Level (SPL), L_{ZTE} : ZTE SPL, L_{STD} : Non-ZTE comparison sequence SPL, **IR**: Inversion Recovery, **SPGR**: Spoiled Gradient Echo, **SE**: Spin Echo, **MPRAGE**: Magnetization Prepared Rapid Gradient Echo, **VFA**: Variable Flip Angle, **MRA**: Magnetic Resonance Angiography, **TOF**: Time of Flight, **DWI**: Diffusion Weighted Imaging, **QSM**: Quantitative Susceptibility Mapping, **mGRE**: Multi Echo Gradient Echo, **GRE-EPI**: Gradient Echo EPI, **N/A**: Not available. Microphone placements: *outside the bore, **inside the bore. Study specific notes: [†] 1 mm protocol, ^{††} 0.8 mm protocol.

very short T_2 , such as cortical bone [76], or myelin [77]. The TR is very short in ZTE, as it is determined only by the readout duration, and the flip angles that can be achieved are typically limited to just above the Ernst angle, unless a very low imaging bandwidth is used [67]. Hence Eq. (4) can be simplified through a first order approximation (i.e., $TR \ll T_1$, and $\alpha \ll 1$ rad) as

$$M_{z,ZTE} \approx \frac{PD}{1 + \frac{T_1}{TR} \cdot \frac{\alpha^2}{2}} \quad (5)$$

Native ZTE image contrast is therefore typically PD weighted with some T_1 saturation. To increase contrast, practical implementations of ZTE typically split up the acquisition into segments to allow for contrast preparation (see Section 3.1), with the gradients being ramped slowly before and after each segment to avoid acoustic noise.

2.4. Image artefacts

Imaging with ZTE has several advantages besides being silent. The low gradient switching rate in ZTE, which makes the acquisition silent, also reduces eddy currents. With a $TE = 0$, there is no time for phase accumulation before the readout, resulting in reduced flow and motion artefacts [45]. However, ZTE is still vulnerable to phase accumulation during the readout, originating from off-resonance effects (e.g., main magnetic field inhomogeneity, tissue susceptibility and fat–water chemical shift). For clinical evaluation of ZTE images, the differences in appearance of off-resonance artefacts, compared to Cartesian acquisitions, are important to consider, as the artefacts otherwise could be misinterpreted as pathology [78]. Chemical shift off-resonance effects can be addressed by using a pixel bandwidth larger than the fat–water chemical shift (i.e., 430 Hz at 3T), and additionally mitigated using a k-space based in-phase and out-of-phase ZTE image decomposition as described by Engström et al [79]. Alternatively, fat saturation pulses can also be used [64,80].

A unique feature of ZTE imaging is the dead-time gap, as discussed in Section 2.1, which results in a spherical region in the centre of k-space without acquired valid samples. Since the centre of k-space encodes low spatial frequencies, the resulting artefact manifests in the form of a slowly varying background signal, rolling off towards the edges of the image (see Fig. 3). The effect is strongest in the centre of the image, and most apparent in areas with

image intensity close to zero, such as the lateral ventricles (yellow arrows), the sinuses (red arrows) and the background.

Radial k-space sampling is in general less sensitive to motion during data acquisition due to the repeated sampling, and hence averaging, of the k-space centre. Motion artefacts therefore appear as localized blurring and streaking instead of coherent ghosting across the imaging FOV seen in Cartesian acquisitions [81]. An example of head motion during data acquisition, comparing ZTE and a Cartesian sampled SPGR, is shown in Fig. 4.

3. ZTE pulse sequences and control of image contrast

The previous section described the basic ZTE pulse sequence and its native contrast behaviour. Because of its low flip angle RF excitation and effective $TE = 0$, the native PD and T_1 -weighted SPGR signal response contains minimal contamination from T_2^* relaxation, susceptibility artefacts, diffusion, and flow effects (see Eq. (4) and (5)). However, to enable use of ZTE in clinical settings, additional contrasts beyond PD and T_1 are required. The following section describes modifications of the ZTE pulse sequence to encode additional contrasts via magnetization preparation and gradient echo refocusing.

3.1. Magnetization prepared ZTE

Magnetization prepared FLASH, as originally described by Haase [83], forms a powerful method to extend the contrast range of SPGR-type sequences beyond native PD and T_1 weighting. For this, the acquisition is divided into segments, each segment starting with a magnetization preparation (MP) module to modify the longitudinal magnetization to contain a desired contrast weighting (e.g., T_1 , T_2 , MT, or diffusion), followed by a certain number of low-FA, short-TR SPGR acquisitions. MP-ZTE offers additional advantages in terms of being silent, $TE = 0$, and fast scanning with short TR where most of the time is used for image encoding. In a segmented ZTE acquisition, the longitudinal magnetization of spoke i within a segment can be expressed as [84]

$$M_{z,i} = \underbrace{M_{prep} \cdot \beta^i}_{Decay} + \underbrace{M_{z,SPGR} \cdot [1 - \beta^i]}_{Recovery}, \quad \text{with } \beta$$

$$= E_1 \cos \alpha, \quad E_1 = e^{-\frac{TR}{T_1}} \quad \forall i = \{0 \dots N-1\} \quad (6)$$




Fig. 3. Example of a dataset reconstructed with and without WASPI to fill the centre of k-space. A clear low spatial frequency artefact appears across the image without WASPI, resulting in erroneous image contrast, especially visible in the lateral ventricles (yellow arrows) and the sinuses (red arrows). Data were acquired with a readout bandwidth of ± 31.25 kHz.




Fig. 4. An example of how motion artefacts manifest as blurring and streaking in ZTE, while in Cartesian SPGR they produce ghosting in the phase-encode direction. Reproduced with permission from Ref. [82].

Here M_{prep} is the longitudinal magnetization at the beginning of the segment, which is weighted by the MP and potential T_1 weighting from incomplete recovery between segments. Eq. (6) shows that the information from the MP is encoded only in the decay term, where it is modulated by T_1 relaxation. For large numbers of spokes (and/or high FA and long TR), $M_{z,i}$ converges towards the steady-state SPGR signal given by Eq. (4) and the decay term (containing the MP weighting) vanishes.

In a Cartesian MP-SPGR acquisition, the first k-space line of each segment can be acquired close to the centre of k-space, i.e., centric ordering, such that the MP weighting dominates the image contrast [85]. In a ZTE acquisition where all readouts, i.e., spokes, originate at the centre of k-space, the image contrast is given by the average signal acquired by all N spokes in a segment, obtained by evaluating the geometric sum of Eq. (6) which yields [67]

$$M_z^{seg} = \underbrace{M_{prep} \cdot f(N, \beta)}_{\text{Decay}} + \underbrace{M_{z,SPGR} \cdot [1 - f(N, \beta)]}_{\text{Recovery}} \quad (7)$$

$$f(N, \beta) = \frac{1}{N} \cdot \frac{1 - \beta^N}{(1 - \beta)}, \quad \text{with } \beta = E_1 \cos \alpha$$

For a single spoke ($N = 1$), $f = 1$ and the acquired signal is equal to the prepared magnetization. For a large number of spokes (i.e., $N \rightarrow \infty$), $f \rightarrow 1$ and the acquired signal converges towards the steady state magnetization $M_{z,SPGR}$.

There are several methods to reduce the influence of undesired T_1 weighting in MP-ZTE. Reducing the number of spokes per readout segment will minimize T_1 weighting from the recovery term in Eq. (6), but at the expense of increased acquisition time as it would increase the number of preparation periods required. Hsu and Lowe proposed a method called *eliminative averaging*, achieved by combining two volumes where the prepared magnetization is $M_{prep,1} = M_{prep}$ and $M_{prep,2} = -M_{prep}$ [84], effectively removing the recovery term and preserving the prepared magnetization. T_1 weighting in the decay term in Eq. (6) can be corrected using a single T_1 assigned to the whole object for a global correction of the k-space data [84]. An alternative method is to apply a k-space filter which increases the relative weighting in the centre of k-space to spokes acquired early in the segment, thus increasing the contribution to the main contrast from these spokes [86], similar to a Cartesian centre phase-encode ordering scheme.




Fig. 5. (A) Gradient waveform structure of a Looping Star sequence with 8 spokes per loop (N_{SPL}) and 2 loops (N_{Loops}), showing the RF and data acquisition (DAQ) scheme for the original (Org.) and coherence resolved (CR) versions of Looping Star. (B) and (C) illustrates the spin coherences for the two versions of Looping Star at four different timepoints during the sequence. The grey shaded region indicates nominal field of view in k-space as defined by the desired image resolution; coherences outside this region (faded arrows) are considered to be dephased and not contributing to the image.

In the applications outlined in Section 4, different types of MP will be described, which produce different contrasts, but all using the same concept of MP followed by one or more ZTE segments. Unless the preparation module includes strong gradients, as for diffusion preparation, the acoustic noise is not increased by MP, and is in fact typically reduced even further, as many MP methods require a delay period for T_1 recovery.

3.2. Multi-echo gradient refocused ZTE

Analogous to multi-echo GRE or UTE, it is possible to design a multi-echo ZTE sequence by refocusing the excited FIDs to produce a gradient echo with T_2^* contrast. Conventionally, bipolar gradients are used for signal refocusing in gradient echo acquisitions, which results in loud acoustic noise from rapid gradient switching. Quiet gradient refocusing with minimal gradient switching can be achieved, as demonstrated with the Looping Star pulse sequence [71] where multiple FID signals are generated and gradient refocused in a looping, time-multiplexed, manner. Similarly, multi-echo ZTE can also be achieved using the BURST technique, albeit at the expense of higher acoustic noise [70]. In this section we will first describe the Looping Star pulse sequence followed by a brief description of ZTE-BURST.

3.2.1. Looping Star

The Looping Star pulse sequence uses a time-multiplexed gradient refocusing scheme to produce T_2^* -weighted gradient echoes, as shown in Fig. 5 [71]. The gradient amplitude is updated directly between spokes, to ensure quiet operation, similar to RUFIS. The k-space trajectory is designed such that each coherence follows a looping trajectory in k-space and periodically refocuses to the centre of k-space to form equidistant gradient echoes.

The acquisition is divided into segments, where each segment encodes a plane in k-space. Multiple segments, rotated relative to each other, are acquired for full 3D k-space coverage. In the first loop of a segment (Fig. 5A), a number of coherences are produced by gradients in different directions, each of which encodes a radial centre-out FID spoke, similar to standard ZTE. In the second loop, the same gradient waveform is then repeated, but without application of RF pulses, to refocus the magnetization and produce T_2^* -weighted gradient echoes (Fig. 5B). Since the magnetization is refocused in a continuous loop, the change in gradient direction and thus acoustic noise can be kept small. The number of spokes acquired per loop (N_{SPL}), together with the duration of each spoke (T_G), governs the acoustic noise and also determines the echo time (TE_{LS}) according to

$$TE_{LS} = N_{SPL} \cdot T_G \quad (8)$$




Fig. 6. (A) Pulse sequence diagram of a single echo gradient refocused ZTE-BURST sequence, with the TE for each coherence indicated by the correspondingly coloured arrow. (B) Visualization of the evolution of the four coherences through the first and second train. The grey circle illustrates the nominal coverage in k-space as determined by the image resolution. K-space sampling is here illustrated in 2D for simplicity, while in practice it is performed in 3D.

The overall acquisition time for a volume (T_{acq}), equivalent to the TR in an fMRI experiment, is given by

$$T_{acq} = (T_G \cdot N_{SPL} \cdot N_{Loops} + 2 \cdot T_{ramp}) \cdot N_{Seg} \quad (9)$$

where N_{Loops} is the number of loops (i.e., FID plus number of echoes), and N_{Seg} is the number of segments which is determined by the level of undersampling in the acquisition. Gradient ramp-up time (T_{ramp}) is typically between 2 and 5 ms to ensure silent operation.

The original version of Looping Star shown in Fig. 5A as Org. LS, corresponding to that initially published by Wiesinger et al., suffers from mixing of the inwards refocusing and outwards dephasing coherences, known as echo-in/echo-out mixing [71] (see Fig. 5B). These signals can be separated through k-space filtering or RF phase cycling, but at the expense of reduced image quality, or increased scan time. In a further development of the pulse sequence, this temporal overlap problem was resolved via separation in time by performing RF excitation only every other spoke, hereafter referred to as coherence resolved Looping Star (CR LS), as shown in Fig. 5A and Fig. 5C. This reduces the number of excited coherences by half, and results in refocusing of spokes with twice the length of the FID, with a piecewise linear trajectory, crossing the centre of k-space in the middle of the readout (see orange arrows in Fig. 5C $t = 8 \dots 10$). The gradient echo spokes are thus sampling along a “curved diameter” without overlap between coherences in the nominal k-space sphere (compare Fig. 5B and C at $t = 8 \dots 10$).

3.2.2. ZTE-BURST

The BURST pulse sequence consists of a series of short, low flip angle, RF pulses applied in the presence of a gradient, which can subsequently be refocused using gradients or refocusing RF pulses (analogous to gradient-echo and spin-echo refocusing techniques) to produce a series of gradient or spin echo signals [87]. Schulte et al. combined the concept of BURST imaging with ZTE [70], as shown in Fig. 6. The sequence consists of trains of N_{SPT} spokes

where in the first train, RF pulses are applied to produce N_{SPT} FIDs, which are encoded separately. In the second train, the RF is turned off and the trajectory is rewound, producing N_{SPT} gradient echoes as the FIDs are refocused, resulting in T_2^* weighting that varies between echoes.

The acoustic noise produced by ZTE-BURST is slightly higher than Looping Star due to rapid gradient switching between the trains. Schulte et al. measured 75.8–78.2 dBA for the ZTE-BURST sequence (depending on the settings), compared to a background scan room level of 66.6 dBA [70] (see Table 1).

4. ZTE for structural neuroimaging

4.1. T_1 contrast mechanisms

Native ZTE imaging provides SPGR-type PD and T_1 contrast weighting, as shown in Eq. (5). Using variable flip angle (VFA) imaging, this can be extended to quantitative PD and T_1 mapping [88]. For a given TR, the degree of T_1 weighting is limited by the maximum achievable flip angle, which is determined by $B_{1,max}$ and by the maximum RF pulse width (τ_{TX}). To avoid slice profile effects, τ_{TX} is limited by the imaging bandwidth (BW_{RX}), as described in Eq. (2). Within these constraints, Ljungberg et al. demonstrated VFA T_1 mapping with RUFIS at 3T using low readout bandwidth (± 7.8 kHz) to achieve high enough flip angles, in the case of that study 12° [67]. Their VFA ZTE acquisition produced T_1 values very similar to standard SPGR acquisition, with equivalent reproducibility and repeatability.

Preliminary results have also demonstrated the utility of VFA T_1 mapping with RUFIS across field strengths [89,90]. As T_1 gets longer with increasing field strength [91], lower flip angles are required to obtain the same T_1 contrast. However, higher field strengths typically require higher readout bandwidth to minimise chemical shift artefacts, and the maximum flip angle is thus reduced. In addition, increased B_1 inhomogeneity at high field will




Fig. 7. Comparison of anatomical T_1 -weighted imaging between Cartesian IR-SPGR using the GE BRAVO (Brain Volume imaging) sequence and ZTE acquired using the MP2RAGE formalism. T_1 map is obtained from the ZTE-MP2RAGE acquisition. Acquisition parameters in Ref. [99].

reduce the effective flip angle further. On the other hand, at lower field strengths, VFA ZTE T_1 mapping is better conditioned, as higher flip angles are possible with lower readout bandwidth requirements and better B_1 uniformity.

In the context of VFA T_1 mapping, Ljungberg et al. also demonstrated a B_1 mapping technique using magnetization prepared ZTE [67,92], similar to the preconditioned TurboFLASH method by Chung et al. [93]. Ljungberg et al. used a train of ultrashort hard RF pulses, similar to those used in the ZTE readout, with a total flip angle of α_{prep} followed by a short readout segment and a T_1 recovery period. The magnetization preparation will thus effectively encode B_1^+ as a global scaling in the image intensity, and a B_1^+ map can be calculated from a set of images with different α_{prep} .

An alternative method to obtain T_1 contrast is to use an inversion or saturation pulse followed by a segmented ZTE readout, in analogy to the MPRAGE (Magnetization Prepared Rapid Gradient Echo) method [94]. Several studies have used inversion recovery (IR) prepared ZTE for T_1 -weighted imaging, with comparison to Cartesian IR prepared SPGR at both 3T [62,63,65,66,95] and 7T [64,80,96]. Applications of IR-ZTE at 7T have used interleaved fat saturation for improved image quality [64,80,97,98]. Similarly, ZTE can be adopted for the MP2RAGE formalism [99] to obtain a bias-field corrected image and quantitative T_1 map, either as two separate acquisitions [100], or as a combined acquisition [101]. Fig. 7 shows an example of a ZTE-MP2RAGE acquisition, demonstrating its ability to produce images with excellent contrast between white and grey matter [101]. Published IR-ZTE results appear promising and demonstrate potential to become equivalent to current MPRAGE T_1 -weighted neuroimaging in terms of image contrast and resolution, especially with development of new image reconstruction techniques such as compressed sensing [102] and Deep Learning [103,104]. Fig. 8 shows an example of an IR-ZTE dataset reconstructed with and without Deep Learning denoising

[103], demonstrating how image details are preserved while noise is reduced.

4.2. T_2 contrast mechanisms

T_2 contrast can be obtained with a ZTE sequence in a manner similar to previous work on T_2 -prepared FLASH [105] and MPRAGE [106,107]. A T_2 -preparation module typically consists of a tip-down pulse, putting the magnetization in the transverse plane where it relaxes with T_2 . A number of refocusing pulses are then applied, after which a tip-up pulse is performed, putting the magnetization back along the longitudinal axis with the desired T_2 weighting [105,108]. For improved robustness to B_0 and B_1 inhomogeneity, adiabatic T_2 preparation such as the mBIR4 pulse can also be used [109,110].

To minimize contribution from T_1 saturation and maintain T_2 contrast, a T_1 -recovery period is required after each segment. The segment should preferably be as short as possible to avoid T_1 recovery during the readout, or alternatively eliminative averaging can be used. There are only a few examples of T_2 -weighted ZTE in the literature; most notably T_2 -prepared ZTE for BOLD fMRI at 3T [74] and 7T [111]. In both studies, each volume was encoded with 1024 spokes, separated into two segments with T_2 preparation before each segment and a 500 ms T_1 -recovery period.

A T_2 -prepared ZTE acquisition can be accelerated with multiple interleaved T_2 -preparation pulses, giving a cumulative effect for each T_1 -recovery period [112], similar to a fast spin echo acquisition in which images with multiple TEs are acquired. While the contrast in such images is dependent on T_2 , as desired, there is also significant influence of T_1 which makes the image appearance diverge from pure T_2 contrast, and also makes T_2 mapping challenging unless T_1 in each voxel is known. It is also possible to obtain T_2 weighting without preparation using spin echo ZTE-




Fig. 8. Example of IR-ZTE dataset reconstructed with and without Deep Learning denoising (DL vs. Std). With DL denoising, the image noise is clearly reduced while still maintaining image resolution and sharpness. White arrows indicate signal from the head rest.

BURST, in which a pair of refocusing RF pulses is inserted between readout trains [70]. Since ZTE-BURST uses very short readout segments, the influence of T_1 is small and so the sequence can be directly used for T_2 mapping [70].

4.3. Multiparametric ZTE

By combining T_1 - and T_2 -preparation modules, T_1 , T_2 and PD can be quantified simultaneously. The approach proposed by Wiesinger et al. [113,114] resembles the method proposed by Kvernby et al. for combined T_1 , T_2 , and PD mapping for cardiac and brain applications [115,116]. In the multi-parametric ZTE method of Wiesinger et al. [113,114], an inversion pulse is applied and several ZTE segments are collected to sample T_1 contrast. As the steady state is approached, T_2 preparation is applied and another ZTE segment, with combined T_1 and T_2 contrast, is acquired. The signal evolution during this transient acquisition can be calculated using the framework outlined in Section 3.1, and the quantitative values can be obtained either through curve fitting or dictionary methods.

Fig. 9 shows an example of quantitative T_1 , T_2 , and PD maps obtained with the multi-parametric ZTE sequence using three ZTE segments after the inversion pulse, followed by a T_2 -preparation module with $TE = 80$ ms [114]. A PD weighted volume was acquired separately and used in the fitting to improve the quantification. Synthetic phase sensitive IR (psIR) and T_2 -weighted fluid attenuated IR (FLAIR) images shown in Fig. 9 were calculated from the PD, T_1 and T_2 maps using an analytic signal equation, evaluated for each voxel. The data shown in Fig. 9 were collected using twofold radial oversampling, as commonly used in ZTE and discussed in Section 2.2, but cropped to show only the central portion of the resulting image. Radial oversampling enables reconstruction of a larger FOV than that prescribed, however. Fig. 10 shows the PD data from Fig. 9 reconstructed without this cropping, which results in an image with twice the prescribed FOV within which signal from the head coil, as indicated by the arrows, is clearly visible.

4.4. Diffusion contrast

Diffusion weighting (DW) can be achieved with a preparation module similar to that used for T_2 preparation but with diffusion gradients added around the refocusing pulse. This approach has been previously demonstrated with Cartesian MPAGE [117–119] and more recently using ZTE by Yuan et al. [69], who used sinusoidal diffusion gradients to reduce the acoustic noise. To remove the T_1 saturation from the readout Yuan et al. used eliminative averaging, as proposed by Hsu et al. [84], while to correct for the T_1 relaxation in the decay term in Eq. (6), they applied a global T_1 correction, estimated from a WASPI acquisition. They also performed phase cycling of the tip up pulse to compensate for eddy currents, resulting in a minimum of 4 volumes acquired per diffusion encoding, b-value. The acquisition time per b-value was 3 min for an in vivo brain scan with $1.56 \times 1.56 \times 6$ mm³ resolution, as shown in Fig. 11. Acoustic noise levels were measured to be only 3 dB above ambient (see Table 1).

Diffusion weighted ZTE has reduced off-resonance artefacts compared to EPI, shown with arrows in Fig. 11. Yuan et al. also reported examples of prostate and knee imaging where distortions using DW-ZTE were markedly reduced compared to standard DW-EPI. Similar improvements were seen in a study by Sandberg et al., where 39 paediatric patients underwent extremity MRI to compare DW-ZTE and DW-EPI in bone marrow, muscle and lesions [120]. DW-ZTE provided similar diffusion metrics to DW-EPI but in some cases with improved image quality due to reduced distortions.

4.5. Magnetization transfer

Magnetization transfer (MT) is an effective and time-efficient mechanism for generating strong contrast between white and grey matter by exploiting large amounts of broad-resonance protons in the lipids of myelin [121,122]. MT preparation can also be used to improve contrast in MRA experiments as it suppresses the signal in the tissue more than the signal from blood [123,124].




Fig. 9. Quantitative PD, T_1 and T_2 maps obtained using a multi-parametric ZTE sequence together with synthetic contrast weighted psIR and T_2 FLAIR images. Acquisition parameters: FOV = $20 \times 20 \times 16 \text{ cm}^3$, resolution = $1 \times 1 \times 1 \text{ mm}^3$, TR = 1.8 ms, BW_{RX} = $\pm 31.25 \text{ kHz}$, TE_{T2Prep} = 80 ms, 256 spokes per segment. FA = 3° for multi-parametric part and FA = 1° for PD volume. Total acquisition time was 6:35 min. The T_1 map and T_2 map have been head masked using the PD volume. Abbreviations: **inv PD** – inverse PD, **psIR** – phase sensitive inversion recovery, **FLAIR** – fluid attenuated inversion recovery. Images have been cropped to head coverage. Fig. 10 shows the PD data reconstructed at twice the prescribed field of view. (Images generated with data from Ref [114]).




Fig. 10. Proton density (PD) image from Fig. 9 reconstructed with twice the field of view, i.e., the fully encoded field of view from the twofold radial oversampling. White arrows highlight part of the receive coil only visible when reconstructed at twice the field of view.

Pulsed MT methods, in which an off-resonance preparation pulse is interleaved with a gradient-echo readout of a steady state acquisition [125], can easily be adapted to a segmented readout [126]. They are hence well suited for implementation with ZTE, as first demonstrated by Holmes et al. [127]. Wood et al. incorporated inhomogeneous MT (ihMT) [128] preparation into a ZTE sequence, showing high specificity to white matter in the brain, as illustrated in Fig. 12 [129]. The ihMT effect requires a material that can sustain dipolar order, and can as such be tuned for increased specificity to the properties of the semi-crystalline myelin sheath [130,131].

Grochowski et al. demonstrated use of MT prepared ZTE for anatomical imaging of the optic nerve at 7T using an adiabatic spectral inversion recovery pulse (ASPIR) for fat saturation, applied 1250 Hz off-resonance, which in addition to reducing the signal from fat also introduces MT contrast [132].

4.6. Angiography

Magnetic Resonance Angiography (MRA) contrast can be achieved by combining ZTE imaging with an Arterial Spin Labelling (ASL) preparation module [133]. Since ASL contrast is commonly produced by a series of low flip angle RF pulses, it does not increase the acoustic noise levels. Shang et al. demonstrated ZTE-MRA with less than 4 dB increase from ambient noise levels [68].

There are several potential benefits of using ZTE for MRA, besides reduced acoustic noise. Early work on RUFIS by Madio and Lowe demonstrated that RUFIS can image turbulent flow [45] and is capable of quantifying flow velocities in the presence of stenosis [134]. These results were corroborated in a phantom study, demonstrating better vascular display accuracy in the presence of stenosis with ASL-PETRA than with 2D and 3D Time Of Flight (TOF) angiography [135]. Further support to these results




Fig. 11. DW-ZTE (A–C) and DW-EPI (D) with $b = 600 \text{ s/mm}^2$. (A) and (B) demonstrate the eddy current artefacts, which are eliminated when combined in (C). Arrows highlights areas with distortion artefacts in DW-EPI that were not present in the DW-ZTE data. Reproduced with permission of John Wiley and Sons from Ref. [69], © 2019 International Society for Magnetic Resonance in Medicine.




Fig. 12. Examples of different MT contrasts acquired with a ZTE sequence. The ihMTR and ihMTRinv images show high sensitivity and specificity to myelin (Abbreviations: PDw – Proton Density weighted, T1w – T₁ weighted, MTw – MT weighted, eMTw – enhanced MT weighted, with dual-sided saturation, MTR – Magnetization Transfer Ratio, eMTR – enhanced MT Ratio, ihMTR – inhomogeneous MTR, the difference between eMTR and MTR, ihMTRinv – inverse ihMTR using the T1w image as a reference instead of the PDw image). Reproduced with permission from Ref. [129] licensed under a CC BY 4.0 license. Figure has been cropped; original figure also contains a row with coefficient of variation.




Fig. 13. Example of ZTE-MRA (**B** and **E**) compared to computed tomography angiogram (CTA) (**A** and **D**) and TOF (**C** and **F**). Top row shows volume rendering and bottom row maximum intensity projection. Image shows a stenosis in a 74-year-old male patient. CTA estimated a 34% stenosis, ZTE estimated 32%, while TOF overestimated the stenosis to 72%. Republished with permission of American Society of Neuroradiology from Ref. [68]; permission conveyed through Copyright Clearance Center, Inc.

was provided in an in vivo study by Shang et al. where ZTE-MRA outperformed TOF imaging for assessment of stenosis and aneurysms in cerebrovascular diseases (see Fig. 13) [68]. Fujiwara et al. also demonstrated improved vessel contrast in the carotid artery with ZTE compared to 3D TOF [136].

Another advantage of ZTE-MRA is reduced artefacts around areas of magnetic susceptibility gradients, as demonstrated for DW-ZTE (Fig. 10), which is relevant for imaging around stents and coils in MRA. These artefacts can be minimised by reducing the TE [137], and thus ZTE sequences have an advantage. Several studies have demonstrated improved vessel visualization around stents and coils with ZTE-MRA compared to conventional gradient echo based TOF-MRA [133,138–141].

The clinical utility of ZTE-MRA is still debated, however, with only a small number of studies published to date. Shang et al. performed a study on 68 patients with suspected cerebrovascular disorder, finding higher inter-modality agreement between ZTE-MRA and computed tomography angiography (CTA), than TOF and CTA [68]. However, in a study by Holdsworth et al. with 51 patients where ZTE-MRA was compared to 3D SPGR-TOF, the ZTE images were rated significantly lower across four categories (image blurring and SNR, lesion conspicuity, and diagnostic confidence), with only 48% of the ZTE-MRA, compared to 90% of SPGR-MRA, scans being judged to be of diagnostic quality [95].

4.7. Quantitative susceptibility mapping and T_2^*

Multi-echo gradient refocused ZTE can be used for quantitative susceptibility mapping (QSM) and T_2^* imaging using either Looping Star [71] or ZTE-BURST [70]. Both methods have demonstrated promising single-subject results, though the 0.8 mm resolution Looping Star QSM protocol was quieter (72.6 dBA) than the 1.00 mm ZTE-BURST protocol (75.8 dBA), both with an acquisition time of approximately 11.5 min.

The TE in Looping Star and ZTE-BURST cannot be chosen arbitrarily, as shown in Eq. (8), and the minimum TE may need to be longer than that of a typical Cartesian acquisition in order to maintain silent operation. In the limiting case with two spokes per loop in Looping Star, the sequence essentially uses bi-polar gradients and is therefore no longer silent. The number of spokes per loop required to maintain silent operation depends on the gradient strength, as this will determine the absolute change in gradient amplitude. Wiesinger et al. demonstrated Looping Star QSM acquired with TE = 26.88 ms using 12 spokes per loop (Fig. 14) which resulted in acoustic noise of 72.6 dBA; for comparison, their 3 mm resolution fMRI acquisition using 32 spokes per loop measured 66.9 dBA [71].

As previously mentioned in the context of DW-ZTE, an advantage of multi-echo ZTE compared to conventional gradient refocused acquisitions is reduced geometric distortions from variations in magnetic susceptibility and reduced artefacts from motion and eddy currents. The high SNR FID (TE = 0) image can be used as a distortion-free reference image for spatial normalization of the gradient echoes, as well as an additional data point in quantitative T_2^* mapping and QSM. Furthermore, the gradient echo spokes can detect fluctuations in the B_0 field as they sample the centre of k-space, as demonstrated by Wiesinger et al. [71]. Finally, the interleaved acquisition strategy used in Looping Star also enables reconstruction of motion navigators during the acquisition for retrospective motion correction.

4.8. Ultra-short T_2 : Bone and myelin imaging

ZTE pulse sequences can capture the MR signal from tissues with ultra-short T_2 , less than 1 ms [142], which typically are considered MR invisible [143]. In neuroimaging, ultra-short T_2 tissues of interest mainly include bone and myelin.




Fig. 14. Example of T_2^* and QSM imaging with Looping Star. Reproduced with permission of John Wiley and Sons from Ref. [71] and adapted to highlight the TE in the right panel. © 2018 International Society for Magnetic Resonance in Medicine.

The utility of ZTE for bone imaging and segmentation has been demonstrated in several studies [76,144–146], as exemplified in Fig. 15A. In order to contrast bone against surrounding soft-tissue, a low flip angle PD-weighted ZTE acquisition with minimal T_1 saturation is used. Chemical shift interferences at fat–water

interfaces must be avoided by using a pixel bandwidth larger than the fat–water chemical shift (i.e., 3.5 ppm) [79].

ZTE images can also be converted to pseudo computed tomography (CT) images as required for attenuation correction in combined positron emission tomography (PET) and MRI acquisitions,




Fig. 15. Ultra-short T_2 PD imaging with ZTE. (A) Skull segmentation. Republished with permission of the American Society of Neuroradiology from Ref. [146]; permission conveyed through Copyright Clearance Center, Inc. (B) Generation of pseudo CT images from ZTE in Hounsfield units, compared to acquired CT (Abbreviations: **ZT** – ZTE-derived pseudo-CT, **RTP** – Radiation Therapy Planning, **Res** – Resolution, **HU** – Hounsfield Units). Reproduced with permission of John Wiley and Sons from Ref. [147], © 2018 International Society for Magnetic Resonance in Medicine.




Fig. 16. Direct myelin imaging with ZTE by Weiger et al. Subtraction of two images with different effective TE yields a qualitative image with contrast between white and grey matter. Reproduced with permission of Elsevier from Ref. [77] under a CC BY-NC-ND 4.0 license.

and MR-only radiation therapy (RT) dose planning [147,148] (see Fig. 15B). Inoue et al. also demonstrated that PD weighted ZTE together with MRA can be a useful tool for endoscopic endonasal transsphenoidal surgery planning, as it allows visualisation of internal carotid arteries as well as cortical bone [149].

Imaging of myelin in the central nervous system is of great interest due to its involvement in numerous diseases, as well as in normal development [150]. Direct imaging of myelin has long been considered impossible due to the ultra-short T_2 of myelin [151,152]. Methods have therefore been developed to probe different proxies for myelin such as ihMT [128], or measurement of water trapped within the myelin lipid bilayers by multi-component T_2 mapping [153]. However, Weiger et al. recently demonstrated direct imaging of ultra-short T_2 components in the brain, attributable in large part to myelin, using ZTE on a customized 3T MR system with a readout bandwidth up to 2000 kHz [77]. By subtracting two images acquired with different bandwidths, resulting in effective TEs of 15 and 460 μ s, they produced a qualitative image with high sensitivity to white matter (see Fig. 16). The total acquisition time was \approx 45 min. Jang et al. have carried out a similar study of imaging ultra-short T_2 components in the brain, but using IR prepared ZTE with a unipolar gradient echo on each spoke, meaning that the sequence is no longer silent [154].

4.9. Other ZTE applications

In addition to the applications described above, ZTE has also been explored for other applications such as MR electrical property (EP) tomography [155] and MR thermometry [156]. In EP tomography, conductivity and permittivity maps are estimated from magnitude and phase variations of the RF magnetic field governed by Maxwell's equations [157]. Lee et al. [155] and more recently also Soull   et al. [158] developed an algorithm for EP mapping based on the product of the transmit (B_1^+) and receive (B_1^-) RF fields [159]. For this purpose, high bandwidth, low-flip angle PD-weighted ZTE images can be used as an approximation of $B_1^- \cdot B_1^+$.

The ZTE-VFA method can be extended for rapid measurement of relative temperature changes based on T_1 temperature dependence [156,160]. In addition to temperature monitoring in soft tissue, ZTE also permits assessing temperature changes in bone structures (e.g., skull) which is important for MR-guided thermal therapies such as high-intensity focused ultrasound (HiFU) [161].

5. ZTE for functional neuroimaging

Functional MRI utilises BOLD contrast to study brain function [162]. This contrast can be observed in T_2 - or T_2^* -weighted acquisitions. A standard ZTE acquisition, e.g., RUFIS, has TE = 0, i.e., no T_2 or T_2^* weighting, and hence does not display BOLD contrast. To achieve functional BOLD contrast with a ZTE sequence, the readout can be preceded by a T_2 -preparation module, which has been demonstrated at both 3T [74] and 7T [111]. In the study at 3T, applications included motor and auditory tasks in four volunteers, where T_2 -prepared RUFIS showed lower sensitivity but improved spatial specificity compared to gradient echo EPI and spin echo EPI. The disadvantage of T_2 -prepared ZTE fMRI is increased acquisition time due to the T_2 preparation and T_1 -recovery period between segments.

The Looping Star sequence can produce T_2^* contrast in the steady state and is therefore a more efficient option for ZTE fMRI. Dionisio-Parra et al. demonstrated the use of Looping Star with a single echo acquisition, as commonly used for GRE-EPI, with visual working memory and resting state paradigms [72] (see Fig. 17). Using a similar acquisition, Wiesinger et al. demonstrated sensitivity to a motor paradigm [71]. Multi-echo fMRI is increasingly being employed due to the benefits of combining echoes, such as reducing the impact of physiological noise. Preliminary results of the multi-echo capability of Looping Star have been presented alongside both block-design and event-related auditory paradigms [73,163], demonstrating good sensitivity to these more complex cognitive tasks.

Mangia et al. have shown preliminary human in vivo results using SWIFT for fMRI [164]. Since SWIFT has TE = 0 it does not produce BOLD contrast. Lehto et al. demonstrated that the main contribution to the observed signal changes in SWIFT fMRI is most likely due to increased blood flow during neuronal activity [165]. Functional imaging with SWIFT has mainly been applied in rodent experiments using either deep brain stimulation (DBS) [165,166] or simultaneous EEG [167]. The advantages of SWIFT compared to standard sequences in these situations are twofold. First, susceptibility artefacts are reduced, resulting in better image quality around the electrodes. Secondly, the low gradient switching rate, which ensures silent acquisition, also results in lower induced currents in the DBS and EEG electrodes and leads. This is particularly important for EEG, where the fast gradient switching can distort the EEG signal [168].




Fig. 17. Images from Dionisio-Parra et al. showing second-level results from an N-back working memory task using (a) Looping Star and (b) Gradient Echo (GE) EPI. (c) Results from a paired *t*-test between the two techniques. Reproduced with permission of John Wiley and Sons from Ref. [72], under a CC BY-NC 4.0 license.

6. Discussion

A ZTE pulse sequence is in essence the simplest spatially-encoded MR pulse sequence one could envision. With RF excitation in the presence of the readout gradient, FID readout and minimal gradient switching between excitations, the acquisition produces low acoustic noise. It also enables very short TRs, on the order of 1 ms, and near 100% sampling efficiency. The native contrast in a ZTE sequence is PD or T_1 , making these contrasts particularly well suited for ZTE, including applications such as bone imaging and ZTE-MP2RAGE. Using Looping Star, T_2^* contrast can also be achieved for fMRI and QSM imaging. To obtain additional image contrasts, magnetization preparation techniques must be employed.

Translation of silent ZTE imaging to clinical settings will require the full suite of clinical image contrasts to be available, including PD, T_1 , T_2 and DW. While PD- and T_1 -weighted images are easily obtained, high quality T_2 -weighted scans exhibit some difficulties with ZTE given its FID acquisition nature. Using multi-parametric ZTE it is possible to perform combined PD, T_1 and T_2 mapping, which then can be used for generating contrast weighted images with different T_2 contrast, as shown in Fig. 9. Synthetic imaging has some limitations though, especially noticeable in T_2 FLAIR images as highlighted in previous studies [169,170]. Wang et al. demonstrated a method for MR image synthesis using a deep learning network which was able to remove many of the artefacts commonly seen in synthetic T_2 FLAIR, such as edge enhancement at tissue interfaces [170]. Furthermore, if the end goal is to perform silent T_2 -weighted imaging, spin echo based sequences with smooth gradient waveform should also be considered and included in future evaluations of silent imaging protocols [63,171,172].

The protocol for DW-ZTE by Yuan et al. showed that diffusion contrast can be achieved, but their acquisition suffered from long acquisition times [69]. To enable clinical translation, further

research should investigate options for optimal acquisition strategies for combination of the phase cycling and eliminative averaging required for this approach. Again, an alternative to DW-ZTE could be EPI based sequences with sinusoidal gradients [173] and reduced slew rates [174] which have shown promise to reduce the acoustic noise. For both T_2 - and DW-prepared ZTE, advanced reconstruction methods, such as deep learning as shown in Fig. 8, could be a way to improve image quality and reduce acquisition time, making them more competitive compared to their non-silent equivalents.

In the field of neurodegeneration, imaging methods for studying myelin are of great interest. Recent developments of ZTE sequences have demonstrated the capacity for sensitizing images to the myelin bilayer. The approach by Weiger et al. utilizing image subtraction at different TE_s is an impressive methodological advancement, but requires long acquisition times [77]. With IR preparation, Jang et al. demonstrated a myelin-sensitized protocol in clinically feasible times, but at the expense of increased acoustic noise from the unipolar gradient echo [154]. Both of these ZTE methods produce images with a signal intensity proportional to the observable MR signal from short-relaxation-time species, which in brain tissue is largely attributable to myelin. Thus, while only semi-quantitative, they represent the closest approaches to date to direct myelin imaging. An alternative approach is ihMT weighted ZTE which is an indirect measure of myelin through magnetization transfer. While giving only an indirect measure of myelin, ihMT has shorter acquisition time, maintains silent operation, and has been shown through histological studies to be highly sensitive to myelin [130].

One patient cohort where silent imaging could find numerous applications is in neonatal imaging, where acoustic noise reduction is required both for hearing protection and to enable scanning during natural sleep [18]. Acoustic noise can be reduced actively using modifications to conventional pulse sequences such as lower slew

rate, but also passively using padding inside the scanner in addition to conventional hearing protection [18,175]. In the developing human connectome project (DHCP), a method of ramping up the gradient amplitude to the desired operational level over a number of repetitions of the sequence in the beginning of the acquisition, in their case a period of 5 s, is used to reduce the startle response, thus not disturbing natural sleep [19]. Considering the constraints put on neonatal imaging with regard to acoustic noise, we believe this is a promising area for adoption of silent ZTE neuroimaging techniques.

Silent MRI techniques could also be helpful for *in utero* MRI, where it is not possible to fit the foetus with hearing protection. While several studies have shown that the surrounding maternal tissue is enough to reduce the acoustic noise down to non-harmful levels [176–180], using a truly quiet sequence would minimize any remaining parental anxiety over this issue.

Finally, in the literature, ZTE sequences go by many different names. RUFIS was one of the first silent ZTE sequences with continuous gradients [45], using algebraic reconstruction to recover the dead-time gap, whereas recent implementations of ZTE on clinical scanners typically use WASPI or PETRA. We recommend using the term ZTE for this general category of pulse sequences and, when applicable, specifying which dead-time gap recovery method is used (PETRA, WASPI, HYFI or algebraic reconstruction). This is particularly useful for neuroimaging applications, where the choice of dead-time gap method has less impact on the image quality and characteristics than in imaging of ultra-short T_2 species [54,143]. Unified terminology will also help accelerate the adoption of these imaging techniques in research studies and clinical practice.

7. Conclusions

Silent ZTE sequences show great potential for use in neuroimaging. The most obvious benefit of swapping to such sequences is the large reduction in acoustic noise, which will greatly increase patient comfort, reduce anxiety, improve communication between the radiographer and subject, and enable a wider range of research into auditory conditions. While ZTE sequences are well suited for PD and T_1 contrasts, and the new Looping Star sequence provides T_2^* and susceptibility contrasts, some standard clinical contrasts such as T_2 and diffusion remain challenging. However, rapid progress is being made with these, and a truly silent comprehensive protocol looks likely to be feasible in the near future.

Declaration of Competing Interest

The authors declare the following financial interests/personal relationships which may be considered as potential competing interests: FW and ABS: employees of General Electric Healthcare. GJB receives honoraria from GE Healthcare for teaching. EL, NLD, TCW, DJL, FZ, SCRW: None.

Acknowledgements

This work was supported by the Wellcome/EPSRC Centre for Medical Engineering at King's College London [WT203148/Z/16/Z], General Electric Healthcare, the NIHR-Wellcome Trust King's Clinical Research Facility, and the National Institute for Health Research (NIHR) Biomedical Research Centre at South London and Maudsley NHS Foundation Trust and King's College London. Nikou Damestani is in receipt of a PhD studentship funded by the NIHR Maudsley Biomedical Research Centre. The views expressed are those of the authors and not necessarily those of the NHS, the NIHR or the Department of Health and Social Care. We also thank the MRC and the Wellcome Trust for ongoing sup-

port of our Neuroimaging research activities at the Centre for Neuroimaging Sciences.

We want to acknowledge that the spin coherence diagrams in Figs. 5 and 6 were inspired by Fig. 1 in Ref. [70].

References

- [1] P. Mansfield, P.M. Glover, J. Beaumont, Sound generation in gradient coil structures for MRI, *Magn. Reson. Med.* 39 (1998) 539–550, <https://doi.org/10.1002/mrm.1910390406>.
- [2] M.J. McJury, Acoustic noise levels generated during high field MR imaging, *Clin. Radiol.* 50 (1995) 331–334, [https://doi.org/10.1016/S0009-9260\(05\)83427-0](https://doi.org/10.1016/S0009-9260(05)83427-0).
- [3] C. Jin, H. Li, X. Li, M. Wang, C. Liu, J. Guo, J. Yang, Temporary hearing threshold shift in healthy volunteers with hearing protection caused by acoustic noise exposure during 3-T multisequence MR neuroimaging, *Radiology*. 286 (2018) 602–608, <https://doi.org/10.1148/radiol.2017161622>.
- [4] M. McJury, F.G. Shellock, Auditory Noise Associated With MR Procedures: A Review, *J. Magn. Reson. Imaging*. 12 (2000) 37–45, [https://doi.org/10.1002/1522-2586\(200007\)12:1<37::AID-JMRI5>3.0.CO;2-I](https://doi.org/10.1002/1522-2586(200007)12:1<37::AID-JMRI5>3.0.CO;2-I).
- [5] J.R. Foster, D.A. Hall, A.Q. Summerfield, A.R. Palmer, R.W. Bowtell, Sound-level measurements and calculations of safe noise dosage during EPI at 3 T, *J. Magn. Reson. Imaging*. 12 (2000) 157–163, [https://doi.org/10.1002/1522-2586\(200007\)12:1<157::AID-JMRI17>3.0.CO;2-M](https://doi.org/10.1002/1522-2586(200007)12:1<157::AID-JMRI17>3.0.CO;2-M).
- [6] A. Moelker, P.A. Wielopolski, P.M.T. Pattynama, Relationship between magnetic field strength and magnetic-resonance-related acoustic noise levels, *Magn. Reson. Mater. Phys., Biol. Med.* 16 (2003) 52–55, <https://doi.org/10.1007/s10334-003-0005-9>.
- [7] R. MacKenzie, C. Sims, R.G. Owens, A.K. Dixon, Patients' perceptions of magnetic resonance imaging, *Clin. Radiol.* 50 (1995) 137–143, [https://doi.org/10.1016/S0009-9260\(05\)83042-9](https://doi.org/10.1016/S0009-9260(05)83042-9).
- [8] I.J. Chou, C.R. Tench, P. Gowland, T. Jaspan, R.A. Dineen, N. Evangelou, R. Abdel-Fahim, W.P. Whitehouse, C.S. Constantinescu, Subjective discomfort in children receiving 3 T MRI and experienced adults' perspective on children's tolerability of 7 T: A cross-sectional questionnaire survey, *BMJ Open*. 4 (2014), <https://doi.org/10.1136/bmjopen-2014-006094> e006094.
- [9] M.J. Versluis, W.M. Teeuwisse, H.E. Kan, M.A. Van Buchem, A.G. Webb, M.J. Van Osch, Subject tolerance of 7 T MRI examinations, *J. Magn. Reson. Imaging*. 38 (2013) 722–725, <https://doi.org/10.1002/jmri.23904>.
- [10] C. Heilmair, J.M. Theysohn, S. Maderwald, O. Kraff, M.E. Ladd, S.C. Ladd, A large-scale study on subjective perception of discomfort during 7 and 1.5T MRI examinations, *Bioelectromagnetics*. 32 (2011) 610–619, <https://doi.org/10.1002/bem.20680>.
- [11] M. Cosottini, D. Frosini, L. Biagi, I. Pesaresi, M. Costagli, G. Tiberi, M. Symms, M. Tosetti, Short-term side-effects of brain MR examination at 7 T: A single-centre experience, *Eur. Radiol.* 24 (2014) 1923–1928, <https://doi.org/10.1007/s00330-014-3177-y>.
- [12] E. Sartoretti, T. Sartoretti, M. Wyss, L. van Smoorenburg, B. Eichenberger, S. van der Duim, D. Cereghetti, C.A. Binkert, S. Sartoretti-Schefer, A. Najafi, Impact of Acoustic Noise Reduction on Patient Experience in Routine Clinical Magnetic Resonance Imaging, *Acad. Radiol.* (2020) 1–8, <https://doi.org/10.1016/j.acra.2020.10.012>.
- [13] M. Toivonen, R. Pääkkönen, S. Savolainen, K. Lehtomäki, Noise attenuation and proper insertion of earplugs into ear canals, *Ann. Occup. Hyg.* 46 (2002) 527–530, <https://doi.org/10.1093/annhyg/mef065>.
- [14] R.E. Brummet, J.M. Talbot, P. Charuhas, Potential Hearing Loss Resulting from MR Imaging, *Magn. Reson. Imaging*. 169 (1988) 539–540, <https://doi.org/10.1148/radiology.169.2.3175004>.
- [15] W. Wagner, I. Staud, G. Frank, F. Dammann, S. Plontke, P.K. Plinkert, Noise in magnetic resonance imaging: No risk for sensorineural function but increased amplitude variability of otoacoustic emissions, *Laryngoscope*. 113 (2003) 1216–1223, <https://doi.org/10.1097/00005537-200307000-00020>.
- [16] E.Y.L. Lim, I.P. Tang, M. Peyman, N. Ramli, P. Narayanan, R. Rajagopalan, 3 Tesla magnetic resonance imaging noise in standard head and neck sequence does not cause temporary threshold shift in high frequency, *Eur. Arch. Oto-Rhino-Laryngol.* 272 (2015) 3109–3113, <https://doi.org/10.1007/s00405-014-3232-y>.
- [17] P. Radomskij, M.A. Schmidt, C.W. Heron, D. Prasher, Effect of MRI noise on cochlear function, *Lancet*. 359 (2002) 1485–1486, [https://doi.org/10.1016/S0140-6736\(02\)08423-4](https://doi.org/10.1016/S0140-6736(02)08423-4).
- [18] D.C. Dean, H. Dirks, J. O'Muircheartaigh, L. Walker, B.A. Jerskey, K. Lehman, M. Han, N. Waskiewicz, S.C.L. Deoni, Pediatric neuroimaging using magnetic resonance imaging during non-sedated sleep, *Pediatr. Radiol.* 44 (2014) 64–72, <https://doi.org/10.1007/s00247-013-2752-8>.
- [19] E.J. Hughes, T. Winchman, F. Padormo, R. Teixeira, J. Wurie, M. Sharma, M. Fox, J. Hutter, L. Cordero-Grande, A.N. Price, J. Allsop, J. Bueno-Conde, N. Tusor, T. Arichi, A.D. Edwards, M.A. Rutherford, S.J. Counsell, J.V. Hajnal, A dedicated neonatal brain imaging system, *Magn. Reson. Med.* 78 (2017) 794–804, <https://doi.org/10.1002/mrm.26462>.
- [20] R.S. Tyler, M. Pienkowski, E.R. Roncancio, H.J. Jun, T. Brozoski, N. Dauman, C.B. Coelho, G. Anderson, A.J. Keiner, A.T. Cacace, N. Martin, B.C.J. More, A Review of Hyperacusis and Future Directions: Part I. Definitions and Manifestations, *Am. J. Audiol.* 23 (2014) 402–419, https://doi.org/10.1044/2014_AJA-14-0010.

- [21] M. Schecklmann, M. Landgrebe, B. Langguth, the TRI Database Study Group, Phenotypic characteristics of hyperacusis in tinnitus, *PLoS One*. 9 (2014), <https://doi.org/10.1371/journal.pone.0086944> e86944.
- [22] L.N. Stiegler, R. Davis, Understanding sound sensitivity in individuals with autism spectrum disorders, *Focus Autism Other Dev. Disabl.* 25 (2010) 67–75, <https://doi.org/10.1177/1088357610364530>.
- [23] D. Tomasi, E.C. Caparelli, L. Chang, T. Ernst, fMRI-acoustic noise alters brain activation during working memory tasks, *Neuroimage*. 27 (2005) 377–386, <https://doi.org/10.1016/j.neuroimage.2005.04.010>.
- [24] T.M. Talavage, W.B. Edmister, P.J. Ledden, R.M. Weisskoff, Quantitative assessment of auditory cortex responses induced by imager acoustic noise, *Hum. Brain Mapp.* 7 (1999) 79–88, [https://doi.org/10.1002/\(SICI\)1097-0193\(1999\)7:2<79::AID-HBM1>3.0.CO;2-R](https://doi.org/10.1002/(SICI)1097-0193(1999)7:2<79::AID-HBM1>3.0.CO;2-R).
- [25] Z.-H. Cho, S.-C. Chung, D.-W. Lim, E.K. Wong, Effects of the acoustic noise of the gradient systems on fMRI: A study on auditory, motor, and visual cortices, *Magn. Reson. Med.* 39 (1998) 331–335, <https://doi.org/10.1002/mrm.1910390224>.
- [26] E.W. Healy, D.C. Moser, K.L. Morrow-Odom, D.A. Hall, J. Fridriksson, Speech Perception in MRI Scanner Noise by Persons With Aphasia, *J. Speech, Lang. Hear. Res.* 50 (2007) 323–334, [https://doi.org/10.1044/1092-4388\(2007\)023](https://doi.org/10.1044/1092-4388(2007)023).
- [27] A.V. Fjaldstad, H.J. Nørgaard, H.M. Fernandes, The Impact of Acoustic fMRI-Noise on Olfactory Sensitivity and Perception, *Neuroscience*. 406 (2019) 262–267, <https://doi.org/10.1016/j.neuroscience.2019.03.028>.
- [28] N. Yakunina, T.S. Kim, W.S. Tae, S.S. Kim, E.C. Nam, Applicability of the sparse temporal acquisition technique in resting-state brain network analysis, *Am. J. Neuroradiol.* 37 (2016) 515–520, <https://doi.org/10.3174/ajnr.A4554>.
- [29] J. Andoh, M. Ferreira, I.R. Leppert, R. Matsushita, B. Pike, R.J. Zatorre, How restless is it with all that noise? Comparison of Interleaved silent steady state (ISSS) and conventional imaging in resting-state fMRI, *Neuroimage*. 147 (2017) 726–735, <https://doi.org/10.1016/j.neuroimage.2016.11.065>.
- [30] M. Smits, S. Kovacs, D. de Ridder, R.R. Peeters, P. van Hecke, S. Sunaert, Lateralization of functional magnetic resonance imaging (fMRI) activation in the auditory pathway of patients with lateralized tinnitus, *Neuroradiology*. 49 (2007) 669–679, <https://doi.org/10.1007/s00234-007-0231-3>.
- [31] D.J. Levitin, V. Menon, J.E. Schmitt, S. Eliez, C.D. White, G.H. Glover, J. Kadis, J. R. Korenberg, U. Bellugi, A.L. Reiss, Neural correlates of auditory perception in Williams syndrome: An fMRI study, *Neuroimage*. 18 (2003) 74–82, <https://doi.org/10.1006/nimg.2002.1297>.
- [32] W.A. Edelstein, R.A. Hedeon, R.P. Mallozzi, S.-A. El-Hamamsy, R.A. Ackermann, T.J. Havens, Making MRI quieter, *Magn. Reson. Imaging*. 20 (2002) 155–163, [https://doi.org/10.1016/S0730-725X\(02\)00475-7](https://doi.org/10.1016/S0730-725X(02)00475-7).
- [33] W.A. Edelstein, T.K. Kidane, V. Taracila, T.N. Baig, T.P. Eagan, Y.-C.N. Cheng, R. W. Brown, J.A. Mallick, Active-passive gradient shielding for MRI acoustic noise reduction, *Magn. Reson. Med.* 53 (2005) 1013–1017, <https://doi.org/10.1002/mrm.20472>.
- [34] N.B. Roozen, A.H. Koevoets, A.J. den Hamer, Active vibration control of gradient coils to reduce acoustic noise of MRI systems, *IEEE/ASME Trans. Mechatronics*. 13 (2008) 325–334, <https://doi.org/10.1109/TMECH.2008.924111>.
- [35] F. Hennel, F. Girard, T. Loenneker, "Silent" MRI with soft gradient pulses, *Magn. Reson. Med.* 42 (1999) 6–10, [https://doi.org/10.1002/\(SICI\)1522-2594\(199907\)42:1<6::AID-MRM2>3.0.CO;2-D](https://doi.org/10.1002/(SICI)1522-2594(199907)42:1<6::AID-MRM2>3.0.CO;2-D).
- [36] F. Hennel, Fast spin echo and fast gradient echo MRI with low acoustic noise, *J. Magn. Reson. Imaging*. 13 (2001) 960–966, <https://doi.org/10.1002/jmri.1138>.
- [37] M. Weiger, D.O. Brunner, B.E. Dietrich, C.F. Muller, K.P. Pruessmann, ZTE Imaging in Humans, *Magn. Reson. Med.* 70 (2013) 328–332, <https://doi.org/10.1002/mrm.24816>.
- [38] M. Weiger, K.P. Pruessmann, MRI with Zero Echo Time, *EMagRes.* 1 (2012) 311–322, <https://doi.org/10.1002/9780470034590.emrstm1292>.
- [39] C.J. Bergin, J.M. Pauly, A. Macovski, Lung parenchyma: Projection reconstruction MR imaging, *Radiology*. 179 (1991) 777–781, <https://doi.org/10.1148/radiology.179.3.2027991>.
- [40] D.J. Tyler, M.D. Robson, R.M. Henkelman, I.R. Young, G.M. Bydder, Magnetic Resonance Imaging With Ultrashort TE (UTE) PULSE Sequences: Technical Considerations, *J. Magn. Reson. Imaging*. 25 (2007) 279–289, <https://doi.org/10.1002/jmri.20851>.
- [41] P. Latta, Z. Starčuk Jr., M.L.H. Gruwel, M.H. Weber, B. Tomanek, K-space trajectory mapping and its application for ultrashort Echo time imaging, *Magn. Reson. Imaging*. 36 (2017) 68–76, <https://doi.org/10.1016/j.mri.2016.10.012>.
- [42] P.M. Margosian, T. Takahashi, M. Takizawa, Practical Implementation of UTE Imaging, *EMagRes.* 1 (2012), <https://doi.org/10.1002/9780470034590.emrstm1247>.
- [43] K.-H. Herrmann, M. Krämer, J.R. Reichenbach, Time efficient 3D radial UTE sampling with fully automatic delay compensation on a clinical 3T MR scanner, *PLoS One*. 11 (2016), <https://doi.org/10.1371/journal.pone.0150371> e0150371.
- [44] S. Hafner, Fast imaging in liquids and solids with the Back-projection Low Angle ShOT (BLAST) technique, *Magn. Reson. Imaging*. 12 (1994) 1047–1051, [https://doi.org/10.1016/0730-725X\(94\)91236-P](https://doi.org/10.1016/0730-725X(94)91236-P).
- [45] D.P. Madio, I.J. Lowe, Ultra-Fast Imaging Using Low Flip Angles and FIDs, *Magn. Reson. Med.* 34 (1995) 525–529, <https://doi.org/10.1002/mrm.1910340407>.
- [46] D. Idayatullin, C. Corum, J.-Y. Park, M. Garwood, Fast and quiet MRI using a swept radiofrequency, *J. Magn. Reson.* 181 (2006) 342–349, <https://doi.org/10.1016/j.jmr.2006.05.014>.
- [47] M. Weiger, K.P. Pruessmann, F. Hennel, MRI with zero echo time: Hard versus sweep pulse excitation, *Magn. Reson. Med.* 66 (2011) 379–389, <https://doi.org/10.1002/mrm.22799>.
- [48] D.M. Grodzki, P.M. Jakob, B. Heismann, Correcting slice selectivity in hard pulse sequences, *J. Magn. Reson.* 214 (2012) 61–67, <https://doi.org/10.1016/j.jmr.2011.10.005>.
- [49] D.O. Kuethe, A. Caprihan, I.J. Lowe, D.P. Madio, H.M. Gach, Transforming NMR data despite missing points, *J. Magn. Reson.* 139 (1999) 18–25, <https://doi.org/10.1006/jmr.1999.1767>.
- [50] K. Schieban, M. Weiger, F. Hennel, A. Boss, K.P. Pruessmann, ZTE imaging with enhanced flip angle using modulated excitation, *Magn. Reson. Med.* 74 (2015) 684–693, <https://doi.org/10.1002/mrm.25464>.
- [51] J. Zhang, D. Idayatullin, C.A. Corum, N. Kobayashi, M. Garwood, Gradient-Modulated SWIFT, *Magn. Reson. Med.* 75 (2016) 537–546, <https://doi.org/10.1002/mrm.25595>.
- [52] N. Kobayashi, U. Goerke, L. Wang, J. Ellerman, G.J. Metzger, M. Garwood, Gradient-Modulated PETRA MRI, *Tomography*. 1 (2015) 85–90, <https://doi.org/10.18383/j.tom.2015.00157>.
- [53] H. Jang, C.N. Wiens, A.B. McMillan, Ramped Hybrid Encoding for Improved Ultrashort Echo Time Imaging, *Magn. Reson. Med.* 76 (2016) 814–825, <https://doi.org/10.1002/mrm.25977>.
- [54] R. Froidevaux, M. Weiger, D.O. Brunner, B.E. Dietrich, B.J. Wilm, K.P. Pruessmann, Filling the dead time gap in zero echo time MRI: principles compared, *Magn. Reson. Med.* 79 (2018) 2036–2045, <https://doi.org/10.1002/mrm.26875>.
- [55] J.J. Van Vaals, M.E. Brummer, W.T. Dixon, H.H. Tuithof, H. Engels, R.C. Nelson, B.M. Gerety, J.L. Chezmar, J.A. den Boer, "Keyhole" method for accelerating imaging of contrast agent uptake, *J. Magn. Reson. Imaging*. 3 (1993) 671–675, <https://doi.org/10.1002/jmri.1880030419>.
- [56] Y. Wu, G. Dai, J.L. Ackerman, M.I. Hrovat, M.J. Glimcher, B.D. Snyder, A. Nazarian, D.A. Chesler, Water- and fat-suppressed proton projection MRI (WASPI) of rat femur bone, *Magn. Reson. Med.* 57 (2007) 554–567, <https://doi.org/10.1002/mrm.21174>.
- [57] D.M. Grodzki, P.M. Jakob, B. Heismann, Ultrashort echo time imaging using pointwise encoding time reduction with radial acquisition (PETRA), *Magn. Reson. Med.* 67 (2012) 510–518, <https://doi.org/10.1002/mrm.23017>.
- [58] R. Froidevaux, M. Weiger, M. Rösler, D. Brunner, K.HYFI Pruessmann, Hybrid filling of the dead-time gap for faster zero echo time imaging, *NMR Biomed* (2021) e4493, <https://doi.org/10.1002/nbm.4493>.
- [59] M.A. Bernstein, K.F. King, X.J. Zhou, Handbook of MRI Pulse Sequences, Elsevier Sci. (2004), <https://doi.org/10.1016/b978-0-12-092861-3.x5000-6>.
- [60] M. Weiger, D.O. Brunner, T. Schmid, R. Froidevaux, M.B. Rösler, S. Gross, K.P. Pruessmann, A virtually 1H-free birdcage coil for zero echo time MRI without background signal, *Magn. Reson. Med.* 78 (2017) 399–407, <https://doi.org/10.1002/mrm.26368>.
- [61] S.T.S. Wong, M.S. Roos, A strategy for sampling on a sphere applied to 3D selective RF pulse design, *Magn. Reson. Med.* 32 (1994) 778–784, <https://doi.org/10.1002/mrm.1910320614>.
- [62] S. Alibek, M. Vogel, W. Sun, D. Winkler, C.A. Baker, M. Burke, H. Gloger, Acoustic noise reduction in MRI using Silent Scan: An initial experience, *Diagnostic Interv. Radiol.* 20 (2014) 360–363, <https://doi.org/10.5152/dir.2014.13458>.
- [63] C. Matsuo-Hagiyama, Y. Watanabe, H. Tanaka, H. Takahashi, A. Arisawa, E. Yoshioka, N. Nabatame, S. Nakano, N. Tomiyama, Comparison of silent and conventional MR imaging for the evaluation of myelination in children, *Magn. Reson. Med. Sci.* 16 (2017) 209–216, <https://doi.org/10.2463/mrms.mp.2016-0045>.
- [64] M. Costagli, M.R. Symms, L. Angeli, D.A.C. Kelley, L. Biagi, A. Farnetani, C. Rua, G. Donatelli, G. Tiberi, M. Tosetti, M. Cosottini, Assessment of Silent T1-weighted head imaging at 7 T, *Eur. Radiol.* 26 (2016) 1879–1888, <https://doi.org/10.1007/s00330-015-3954-2>.
- [65] M. Ida, T. Wakayama, M.L. Nielsen, T. Abe, D.M. Grodzki, Quiet T1-weighted imaging using PETRA: Initial clinical evaluation in intracranial tumor patients, *J. Magn. Reson. Imaging*. 41 (2015) 447–453, <https://doi.org/10.1002/jmri.24575>.
- [66] N. Aida, T. Niwa, Y. Fujii, K. Nozawa, M. Enokizono, K. Murata, T. Obata, Quiet T1-Weighted pointwise encoding time reduction with radial acquisition for assessing myelination in the pediatric brain, *Am. J. Neuroradiol.* 37 (2016) 1528–1534, <https://doi.org/10.3174/ajnr.A4747>.
- [67] E. Ljungberg, T. Wood, A.B. Solana, S. Kolind, S.C.R. Williams, F. Wiesinger, G.J. Barker, Silent T1 mapping using the variable flip angle method with B1 correction, *Magn. Reson. Med.* 84 (2020) 813–824, <https://doi.org/10.1002/mrm.28178>.
- [68] S. Shang, J. Ye, W. Dou, X. Luo, J. Qu, Q. Zhu, H. Zhang, J. Wu, Validation of zero TE-MRA in the characterization of cerebrovascular diseases: A feasibility study, *Am. J. Neuroradiol.* 40 (2019) 1484–1490, <https://doi.org/10.3174/ajnr.A6173>.
- [69] J. Yuan, Y. Hu, A. Menini, C.M. Sandino, J. Sandberg, V. Sheth, C.J. Moran, M. Alley, M. Lustig, B. Hargreaves, S. Vasanawala, Near-silent distortionless DWI using magnetization-prepared RUFIS, *Magn. Reson. Med.* 84 (2020) 170–181, <https://doi.org/10.1002/mrm.28106>.

- [70] R.F. Schulte, G. Buonincontri, M. Costagli, A. Menini, F. Wiesinger, A.B. Solana, Silent T2* and T2 encoding using ZTE combined with BURST, *Magn. Reson. Med.* 81 (2019) 2277–2287, <https://doi.org/10.1002/mrm.27552>.
- [71] F. Wiesinger, A. Menini, A.B. Solana, Looping Star, *Magn. Reson. Med.* 81 (2019) 57–68, <https://doi.org/10.1002/mrm.27440>.
- [72] B. Dionisio-Parra, F. Wiesinger, P.G. Sämann, M. Czisch, A.B. Solana, Looping Star fMRI in Cognitive Tasks and Resting State, *J. Magn. Reson. Imaging* 52 (2020) 739–751, <https://doi.org/10.1002/jmri.27073>.
- [73] N.L. Damestani, D.J. Lythgoe, F. Wiesinger, A.B. Solana, S.C.R. Williams, F. Zelaya, Looping Star silent fMRI: a platform for improving studies of auditory processing, in: *Proc. Intl. Soc. Mag. Reson. Med.* 27, Montreal, QC, Canada, 2019: p. 363. <https://archive.ismrm.org/2019/0363.html>.
- [74] A.B. Solana, A. Menini, L.L. Sacolick, N. Hehn, F. Wiesinger, Quiet and Distortion-Free, Whole Brain BOLD fMRI Using T2 -Prepared RUFIS, *Magn. Reson. Med.* 75 (2016) 1402–1412, <https://doi.org/10.1002/mrm.25658>.
- [75] A. Haase, J. Frahm, D. Matthaei, W. Hanicke, K.-D. Merboldt, FLASH imaging, Rapid NMR imaging using low flip-angle pulses, *J. Magn. Reson.* 67 (1986) 258–266, [https://doi.org/10.1016/0022-2364\(86\)90433-6](https://doi.org/10.1016/0022-2364(86)90433-6).
- [76] F. Wiesinger, L.L. Sacolick, A. Menini, S.S. Kaushik, S. Ahn, P. Veit-Haibach, G. Delso, D.D. Shanhag, Zero TE MR bone imaging in the head, *Magn. Reson. Med.* 75 (2016) 107–114, <https://doi.org/10.1002/mrm.25545>.
- [77] M. Weiger, R. Froidevaux, E.L. Baadsvik, D.O. Brunner, M.B. Rosler, K.P. Pruessmann, Advances in MRI of the myelin bilayer, *Neuroimage* 217 (2020), <https://doi.org/10.1016/j.neuroimage.2020.116888>.
- [78] M. Bydder, J. Du, A. Takahashi, A. Shimakawa, G. Hamilton, S. Sinha, G.M. Bydder, Chemical Shift Artifact in Center-Out Radial Sampling: A Potential Pitfall in Clinical Diagnosis, in: *Proc. Intl. Soc. Mag. Reson. Med.* 15, Berlin, 2007: p. 1811. <https://cds.ismrm.org/ismrm-2007/files/01811.pdf>.
- [79] M. Engström, G. McKinnon, C. Cozzini, F. Wiesinger, In-phase zero TE musculoskeletal imaging, *Magn. Reson. Med.* 83 (2020) 195–202, <https://doi.org/10.1002/mrm.27928>.
- [80] P.E.Z. Larson, M. Han, R. Krug, A. Jakary, S.J. Nelson, D.B. Vigneron, R.G. Henry, G. McKinnon, D.A.C. Kelley, Ultrashort echo time and zero echo time MRI at 7T, *Magn. Reson. Mater. Physics, Biol. Med.* 29 (2016) 359–370, <https://doi.org/10.1007/s10334-015-0509-0>.
- [81] G.H. Glover, J.M. Pauly, Projection Reconstruction Techniques for Reduction of Motion Effects in MRI, *Magn. Reson. Med.* 28 (1992) 275–289, <https://doi.org/10.1002/mrm.1910280209>.
- [82] E. Ljungberg, MRI with Zero Echo Time: Quick, Quiet, King's College London, Quantitative, 2020.
- [83] A. Haase, FLASH MRI. Snapshot, Applications to T1, T2, and Chemical-Shift imaging, *Magn. Reson. Med.* 13 (1990) 77–89, <https://doi.org/10.1002/mrm.1910130109>.
- [84] J.-J. Hsu, I.J. Lowe, Spin-lattice relaxation and a fast T1-map acquisition method in MRI with transient-state magnetization, *J. Magn. Reson.* 169 (2004) 270–278, <https://doi.org/10.1016/j.jmr.2004.05.001>.
- [85] A.E. Holsinger, S.J. Riederer, The importance of phase-encoding order in ultra-short TR snapshot MR imaging, *Magn. Reson. Med.* 16 (1990) 481–488, <https://doi.org/10.1002/mrm.1910160315>.
- [86] H.K. Song, L. Dougherty, k-Space weighted image contrast (KWIC) for contrast manipulation in projection reconstruction MRI, *Magn. Reson. Med.* 44 (2000) 825–832, [https://doi.org/10.1002/1522-2594\(200012\)44:6<825::AID-MRM2>3.0.CO;2-D](https://doi.org/10.1002/1522-2594(200012)44:6<825::AID-MRM2>3.0.CO;2-D).
- [87] J. Hennig, M. Hodapp, Burst imaging, *Magma Magn. Reson. Mater. Physics, Biol. Med.* 1 (1993) 39–48, <https://doi.org/10.1007/BF02660372>.
- [88] E.K. Fram, R.J. Herfkens, G.A. Johnson, G.H. Glover, J.P. Karis, A. Shimakawa, T. G. Perkins, N.J. Pelc, Rapid calculation of T1 using variable flip angle gradient refocused imaging, *Magn. Reson. Imaging* 5 (1987) 201–208, [https://doi.org/10.1016/0730-725X\(87\)90021-X](https://doi.org/10.1016/0730-725X(87)90021-X).
- [89] E. Ljungberg, B. Burns, P.E.Z. Larson, S. Kolind, M. Symms, F. Wiesinger, G.J. Barker, Silent T1-Mapping at 7T Using the Variable Flip Angle Method, in: *Proc. Intl. Soc. Mag. Reson. Med.* 27, Montreal, 2019: p. 4976. <https://archive.ismrm.org/2019/4976.html>.
- [90] E. Ljungberg, B. Burns, T. Wood, A.B. Solana, P.E.Z. Larson, G.J. Barker, F. Wiesinger, ZTE Imaging Across Field Strengths: Opportunities for Low-Field Imaging, in: *Proc. Intl. Soc. Mag. Reson. Med.* 28, Virtual, 2020: p. 523. <https://archive.ismrm.org/2020/0523.html>.
- [91] W.D. Rooney, G. Johnson, X. Li, E.R. Cohen, S.G. Kim, K. Ugurbil, C.S. Springer, Magnetic field and tissue dependencies of human brain longitudinal 1H2O relaxation in vivo, *Magn. Reson. Med.* 57 (2007) 308–318, <https://doi.org/10.1002/mrm.21122>.
- [92] E. Ljungberg, F. Wiesinger, A.B. Solana, G.J. Barker, Silent Magnetization Prepared B1-map Acquisition - SIMBA, in: *Proc. Intl. Soc. Mag. Reson. Med.* 27, Montreal, QC, Canada, 2019: p. 0448. <https://archive.ismrm.org/2019/0448.html>.
- [93] S. Chung, D. Kim, E. Breton, L. Axel, Rapid B1+ Mapping Using a Preconditioning RF Pulse with TurboFLASH Readout, *Magn. Reson. Med.* 64 (2010) 439–446, <https://doi.org/10.1002/mrm.22423>.
- [94] J.P. Mugler III, J.R. Brookeman, Three-dimensional magnetization-prepared rapid gradient-echo imaging (3D MP RAGE), *Magn. Reson. Med.* 15 (1990) 152–157, <https://doi.org/10.1002/mrm.1910150117>.
- [95] S.J. Holdsworth, S.J. Macpherson, K.W. Yeom, M. Wintermark, G. Zaharchuk, Clinical evaluation of silent T1-weighted MRI and silent MR angiography of the brain, *Am. J. Roentgenol.* 210 (2018) 404–411, <https://doi.org/10.2214/AJR.17.18247>.
- [96] L. Sacolick, F. Wiesinger, D.A.C. Kelley, M.M. Khalighi, B.K. Rutt, 7 Tesla zero-echo time imaging of the head, in: *ESMRMB 2013, 30th Annu. Sci. Meet. Toulouse, Fr.* 3–5 Oct. Abstr. Thursday, *Magn. Reson. Mater. Phys.* 2013: p. 53, doi: <https://doi.org/10.1007/s10334-013-0382-7>.
- [97] D.A.C. Kelley, A. Jakary, R. Henry, S.J. Nelson, D.B. Vigneron, Depiction of Multiple Sclerosis Lesions with Zero Echo Time (ZTE) Imaging at 7T, in: *Proc. Intl. Soc. Mag. Reson. Med.* 22, Milan, Italy, 2014: p. 2068. <https://cds.ismrm.org/protected/14MPProceedings/files/2068.pdf>.
- [98] D.A.C. Kelley, G.C. McKinnon, L. Sacolick, D.B. Vigneron, S.J. Nelson, Optimization of a Zero Echo Time (ZTE) Sequence at 7T with Phased Array Coils, in: *Proc. Intl. Soc. Mag. Reson. Med.* 22, Milan, Italy, 2014: p. 4933. <https://archive.ismrm.org/2014/4933.html>.
- [99] J.P. Marques, T. Kober, G. Krueger, W. van der Zwaag, P.-F. Van de Moortele, R. Gruetter, MP2RAGE, a self bias-field corrected sequence for improved segmentation and T1-mapping at high field, *Neuroimage* 49 (2010) 1271–1281, <https://doi.org/10.1016/j.neuroimage.2009.10.002>.
- [100] M. Symms, F. Wiesinger, M. Costagli, D. Kelley, M. Cosottini, M. Tosetti, Silent Corrected Using Second Image (SCUSI) - Application of the MP2RAGE formalism to T1-weighted Zero Time Echo Imaging, in: *Proc. Intl. Soc. Mag. Reson. Med.* 26, Paris, 2018: p. 2048. <https://archive.ismrm.org/2018/2048.html>.
- [101] E. Ljungberg, F. Wiesinger, T.C. Wood, Silent Structural Imaging and T1-mapping with a Rapid-Radial Twice-Prepared (R2P2) Sequence, in: *Proc. Intl. Soc. Mag. Reson. Med.* 27, Montreal, 2019: p. 4636. <https://archive.ismrm.org/2019/4636.html>.
- [102] M. Lustig, D. Donoho, J.M. Pauly, Sparse MRI: The application of compressed sensing for rapid MR imaging, *Magn. Reson. Med.* 58 (2007) 1182–1195, <https://doi.org/10.1002/mrm.21391>.
- [103] R.M. Lebel, Performance characterization of a novel deep learning-based MR image reconstruction pipeline, *Arxiv.* (2020), <https://arxiv.org/abs/2008.06559>.
- [104] K. Hammernik, T. Klatzer, E. Kobler, M.P. Recht, D.K. Sodickson, T. Pock, F. Knoll, Learning a variational network for reconstruction of accelerated MRI data, *Magn. Reson. Med.* 79 (2018) 3055–3071, <https://doi.org/10.1002/mrm.26977>.
- [105] R. Deichmann, H. Adolf, U. Nöth, S. Morrissey, C. Schwarzbauer, A. Haase, Fast T2-mapping with SNAPSHOT FLASH Imaging, *Magn. Reson. Imaging* 13 (1995) 633–639, [https://doi.org/10.1016/0730-725X\(95\)00004-Z](https://doi.org/10.1016/0730-725X(95)00004-Z).
- [106] J.P. Mugler III, T.A. Spraggins, J.R. Brookeman, T2-weighted Three-dimensional MP-RAGE MR Imaging, *J. Magn. Reson. Imaging* 1 (1991) 731–737, <https://doi.org/10.1002/jmri.1880010621>.
- [107] J. Conklin, J.D. Winter, R.T. Thompson, N. Gelman, High-contrast 3D neonatal brain imaging with combined T1- and T2-weighted MP-RAGE, *Magn. Reson. Med.* 59 (2008) 1190–1196, <https://doi.org/10.1002/mrm.21548>.
- [108] J.H. Brittain, B.S. Hu, G.A. Wright, C.H. Meyer, A. Macovski, D.G. Nishimura, Coronary Angiography with Magnetization-Prepared T2 Contrast, *Magn. Reson. Med.* 33 (1995) 689–696, <https://doi.org/10.1002/mrm.1910330515>.
- [109] M. Garwood, L. DelaBarre, The return of the frequency sweep: Designing adiabatic pulses for contemporary NMR, *J. Magn. Reson.* 153 (2001) 155–177, <https://doi.org/10.1006/jmre.2001.2340>.
- [110] R. Nezafat, R. Ouwerkerk, A.J. Derbyshire, M. Stuber, E.R. McVeigh, Spectrally selective B1-insensitive T2 magnetization preparation sequence, *Magn. Reson. Med.* 61 (2009) 1326–1335, <https://doi.org/10.1002/mrm.21742>.
- [111] B. Burns, T. Flynn, A.B. Solana, P. Cao, A. Jakary, A. Menini, F. Wiesinger, P. Larson, Quiet, Low-Distortion Whole Brain T2-BOLD fMRI at 7T, in: *Proc. Intl. Soc. Mag. Reson. Med.* 27, Montreal, QC, Canada, 2019: p. 3930. <https://archive.ismrm.org/2019/3930.html>.
- [112] E. Ljungberg, B. Burns, T. Wood, S. Kolind, F. Wiesinger, G.J. Barker, Rapid, multi-TE, T2-prepared RUFIS for Silent T2-weighted imaging, in: *Proc. Intl. Soc. Mag. Reson. Med.* 27, Montreal, 2019: p. 4571. <https://archive.ismrm.org/2019/4571.html>.
- [113] F. Wiesinger, M.A. Janich, E. Ljungberg, G.J. Barker, A.B. Solana, Silent, 3D MR Parameter Mapping using Magnetization Prepared Zero TE, in: *Proc. Intl. Soc. Mag. Reson. Med.* 26, Paris, 2018: p. 0061. <https://archive.ismrm.org/2018/0061.html>.
- [114] F. Wiesinger, E. Ljungberg, M. Engström, S. Kaushik, T. Wood, S. Williams, G. Barker, A.B. Solana, PSST. . . Parameter mapping Swift and Silent, in: *Proc. Intl. Soc. Mag. Reson. Med.* 28, 2020: p. 882. <https://archive.ismrm.org/2020/0882.html>.
- [115] S. Kvernby, M.J.B. Warntjes, H. Haraldsson, C.J. Carlhäll, J. Engvall, T. Ebbens, Simultaneous three-dimensional myocardial T1 and T2 mapping in one breath hold with 3D-QALAS, *J. Cardiovasc. Magn. Reson.* 16 (2014) 102, <https://doi.org/10.1186/s12968-014-0102-0>.
- [116] S. Fujita, A. Hagiwara, M. Hori, M. Warntjes, K. Kamagata, I. Fukunaga, M. Goto, H. Takuya, K. Takasu, C. Andica, T. Maekawa, M.Y. Takemura, R. Irie, A. Wada, M. Suzuki, S. Aoki, 3D quantitative synthetic MRI-derived cortical thickness and subcortical brain volumes: Scan-rescan repeatability and comparison with conventional T1-weighted images, *J. Magn. Reson. Imaging* 50 (2019) 1834–1842, <https://doi.org/10.1002/jmri.26744>.
- [117] H. Lee, R.R. Price, Diffusion imaging with the MP-RAGE sequence, *J. Magn. Reson. Imaging* 4 (1994) 837–842, <https://doi.org/10.1002/jmri.1880040616>.

- [118] T. Numano, K. Homma, T. Hirose, Diffusion-weighted three-dimensional MP-RAGE MR imaging, *Magn. Reson. Imaging*. 23 (2005) 463–468, <https://doi.org/10.1016/j.mri.2004.12.002>.
- [119] T. Numano, K. Homma, N. Iwasaki, K. Hyodo, N. Nitta, T. Hirose, In vivo isotropic 3D diffusion tensor mapping of the rat brain using diffusion-weighted 3D MP-RAGE MRI, *Magn. Reson. Imaging*. 24 (2006) 287–293, <https://doi.org/10.1016/j.mri.2005.12.011>.
- [120] J.K. Sandberg, V.A. Young, A.B. Syed, J. Yuan, Y. Hu, C. Sandino, A. Menini, B. Hargreaves, S. Vasanawala, Near-Silent and Distortion-Free Diffusion MRI in Pediatric Musculoskeletal Disorders: Comparison With Echo Planar Imaging Diffusion, *J. Magn. Reson. Imaging*. 53 (2021) 504–513, <https://doi.org/10.1002/jmri.27330>.
- [121] S.D. Wolff, R.S. Balaban, Magnetization transfer contrast (MTC) and tissue water proton relaxation in vivo, *Magn. Reson. Med.* 10 (1989) 135–144, <https://doi.org/10.1002/mrm.1910100113>.
- [122] R.M. Henkelman, G.J. Stanisz, S.J. Graham, Magnetization transfer in MRI: A review, *NMR Biomed.* 14 (2001) 57–64, <https://doi.org/10.1002/nbm.683>.
- [123] R.R. Edelman, S.S. Ahn, D. Chien, W. Li, A. Goldmann, M. Mantello, J. Kramer, J. Kleefeld, Improved time-of-flight MR angiography of the brain with magnetization transfer contrast, *Radiology*. 184 (1992) 395–399, <https://doi.org/10.1148/radiology.184.2.1620835>.
- [124] G.B. Pike, B.S. Hu, G.H. Glover, D.R. Enzmann, Magnetization transfer time-of-flight magnetic resonance angiography, *Magn. Reson. Med.* 25 (1992) 372–379, <https://doi.org/10.1002/mrm.1910250217>.
- [125] S.J. Graham, R.M. Henkelman, Understanding Pulsed Magnetization Transfer, *J. Magn. Reson. Imaging*. 7 (1997) 903–912, <https://doi.org/10.1002/jmri.1880070520>.
- [126] S. Mchinda, G. Varma, V.H. Prevost, A. Le Troter, S. Rapacchi, M. Guye, J. Pelletier, J.-P. Ranjeva, D.C. Alsop, G. Duhamel, O.M. Girard, Whole brain inhomogeneous magnetization transfer (ihMT) imaging: Sensitivity enhancement within a steady-state gradient echo sequence, *Magn. Reson. Med.* 79 (2018) 2607–2619, <https://doi.org/10.1002/mrm.26907>.
- [127] J.H. Holmes, A. Samsonov, P. Mossahebi, D. Hernando, A.S. Field, K.M. Johnson, Rapid, motion robust, and quiet quantitative magnetization transfer (qMT) imaging using a zero echo time (ZTE) acquisition, in: *Proc. Intl. Soc. Mag. Reson. Med* 23, Toronto, ON, 2015: p. 3358. <https://archive.ismrm.org/2015/3358.html>.
- [128] G. Varma, G. Duhamel, C. de Bazelaire, D.C. Alsop, Magnetization transfer from inhomogeneously broadened lines: A potential marker for myelin, *Magn. Reson. Med.* 73 (2015) 614–622, <https://doi.org/10.1002/mrm.25174>.
- [129] T.C. Wood, N.L. Damestani, A.J. Lawrence, E. Ljungberg, G.J. Barker, A.B. Solana, F. Wiesinger, S.C.R. Williams, Silent myelin-weighted magnetic resonance imaging [version 2; peer review: 3 approved, 1 approved with reservations], *Wellcome Open Res* 5 (2020), <https://doi.org/10.12688/wellcomeopenres.15845.2>.
- [130] G. Duhamel, V.H. Prevost, M. Cayre, A. Hertenan, S. Mchinda, V.N. Carvalho, G. Varma, P. Durbec, D.C. Alsop, O.M. Girard, Validating the sensitivity of inhomogeneous magnetization transfer (ihMT) MRI to myelin with fluorescence microscopy, *Neuroimage*. 199 (2019) 289–303, <https://doi.org/10.1016/j.neuroimage.2019.05.061>.
- [131] A.I. Boulterne, The history of myelin, *Exp. Neurol.* 283 (2016) 431–445, <https://doi.org/10.1016/j.expneurol.2016.06.005>.
- [132] C. Grochowski, M. Symms, K. Jonak, P. Krukow, T. Wood, E. Ljungberg, J. Ensenat, K. Nowomiejska, R. Rejdak, R. Maciejewski, G.J. Barker, The Evaluation of Optic Nerves Using 7 Tesla “Silent” Zero Echo Time Imaging in Patients with Leber’s Hereditary Optic Neuropathy with or without Idebenone Treatment, *J. Clin. Med.* 9 (2020) 1112, <https://doi.org/10.3390/jcm9041112>.
- [133] R. Irie, M. Suzuki, M. Yamamoto, N. Takano, Y. Suga, M. Hori, K. Kamagata, M. Takayama, M. Yoshida, S. Sato, N. Hamasaki, H. Oishi, S. Aoki, Assessing Blood Flow in an Intracranial Stent: A Feasibility Study of MR Angiography Using a Silent Scan after Stent-Assisted Coil Embolization for Anterior Circulation Aneurysms, *Am. J. Neuroradiol.* 36 (2015) 967–970, <https://doi.org/10.3174/ajnr.A4199>.
- [134] D.P. Madio, H.M. Gach, I.J. Lowe, Ultra-fast Velocity Imaging in Stenotically Produced Turbulent Jets Using RUFIS, *Magn. Reson. Med.* 39 (1998), <https://doi.org/10.1002/mrm.1910390410>.
- [135] I. Koktzoglou, S. Giri, D. Piccini, D.M. Grodzki, O. Flanagan, I.G. Murphy, N. Gupta, J.D. Collins, R.R. Edelman, Arterial spin labeled carotid MR angiography: A phantom study examining the impact of technical and hemodynamic factors, *Magn. Reson. Med.* 75 (2016) 295–301, <https://doi.org/10.1002/mrm.25611>.
- [136] Y. Fujiwara, Y. Muranaka, Improvement in visualization of carotid artery uniformity using silent magnetic resonance angiography, *Radiol. Phys. Technol.* 10 (2017) 113–120, <https://doi.org/10.1007/s12194-016-0375-0>.
- [137] S. Kakeda, Y. Korogi, Y. Hiai, T. Sato, N. Ohnari, J. Moriya, K. Kamada, MRA of intracranial aneurysms embolized with platinum coils: A vascular phantom study at 1.5T and 3T, *J. Magn. Reson. Imaging*. 28 (2008) 13–20, <https://doi.org/10.1002/jmri.21421>.
- [138] N. Takano, M. Suzuki, R. Irie, M. Yamamoto, K. Teranishi, K. Yatomi, N. Hamasaki, K.K. Kumamaru, M. Hori, H. Oishi, S. Aoki, Non-contrast-enhanced silent scan MR angiography of intracranial anterior circulation aneurysms treated with a low-profile visualized intraluminal support device, *Am. J. Neuroradiol.* 38 (2017) 1610–1616, <https://doi.org/10.3174/ajnr.A5223>.
- [139] N. Takano, M. Suzuki, R. Irie, M. Yamamoto, N. Hamasaki, K. Kamagata, K.K. Kumamaru, M. Hori, H. Oishi, S. Aoki, Usefulness of non-contrast-enhanced MR angiography using a silent scan for follow-up after Y-configuration stent-assisted coil embolization for basilar tip aneurysms, *Am. J. Neuroradiol.* 38 (2017) 577–581, <https://doi.org/10.3174/ajnr.A5033>.
- [140] S. Shang, J. Ye, X. Luo, J. Qu, Y. Zhen, J. Wu, Follow-up assessment of coiled intracranial aneurysms using zTE MRA as compared with TOF MRA: a preliminary image quality study, *Eur. Radiol.* 27 (2017) 4271–4280, <https://doi.org/10.1007/s00330-017-4794-z>.
- [141] Y.J. Heo, H.W. Jeong, J.W. Baek, S.T. Kim, Y.G. Jeong, J.Y. Lee, S.C. Jin, Pointwise encoding time reduction with radial acquisition with subtraction-based MRA during the follow-up of stent-assisted coil embolization of anterior circulation aneurysms, *Am. J. Neuroradiol.* 40 (2019) 815–819, <https://doi.org/10.3174/ajnr.A6035>.
- [142] G.M. Bydder, The Agfa Mayneord lecture: MRI of short and ultrashort T2 and T2* components of tissues, fluids and materials using clinical systems, *Br. J. Radiol.* 84 (2011) 1067–1082, <https://doi.org/10.1259/bjr/74368403>.
- [143] M. Weiger, K.P. Pruessmann, Short-T2 MRI: Principles and recent advances, *Prog. Nucl. Magn. Reson. Spectrosc.* 114–115 (2019) 237–270, <https://doi.org/10.1016/j.pnmrs.2019.07.001>.
- [144] G. Delso, F. Wiesinger, L.I. Sacolick, S.S. Kaushik, D.D. Shanbhag, M. Hüllner, P. Veit-Haibach, Clinical evaluation of zero-echo-time MR imaging for the segmentation of the skull, *J. Nucl. Med.* 56 (2015) 417–422, <https://doi.org/10.2967/jnumed.114.149997>.
- [145] R.E. Breighner, Y. Endo, G.P. Konin, L.V. Gulotta, M.F. Koff, H.G. Potter, Zero echo time imaging of the shoulder: Enhanced osseous detail by using MR imaging, *Radiology*. 286 (2018) 960–966, <https://doi.org/10.1148/radiol.2017170906>.
- [146] A. Lu, K.R. Gorny, M.-L. Ho, Zero TE MRI for craniofacial bone imaging, *Am. J. Neuroradiol.* 40 (2019) 1562–1566, <https://doi.org/10.3174/ajnr.A6175>.
- [147] F. Wiesinger, M. Bylund, J. Yang, S. Kaushik, D. Shanbhag, S. Ahn, J.H. Jonsson, J.A. Lundman, T. Hope, T. Nyholm, P. Larson, C. Cozzini, Zero TE-based pseudo-CT image conversion in the head and its application in PET/MR attenuation correction and MR-guided radiation therapy planning, *Magn. Reson. Med.* 80 (2018) 1440–1451, <https://doi.org/10.1002/mrm.27134>.
- [148] A.P. Leynes, J. Yang, F. Wiesinger, S.S. Kaushik, D.D. Shanbhag, Y. Seo, T.A. Hope, P.E.Z. Larson, Zero-echo-time and dixon deep pseudo-CT (ZeDD CT): Direct generation of pseudo-CT images for Pelvic PET/MRI Attenuation Correction Using Deep Convolutional Neural Networks with Multiparametric MRI, *J. Nucl. Med.* 59 (2018) 852–858, <https://doi.org/10.2967/jnumed.117.198051>.
- [149] A. Inoue, S. Kohno, N. Nishida, S. Suehiro, S. Matsumoto, M. Nishikawa, S. Ozaki, Y. Nakamura, S. Shigekawa, H. Watanabe, H. Senba, B. Matsuura, T. Ohnishi, T. Kunieda, Clinical utility of new three-dimensional model using a zero-echo-time sequence in endoscopic endonasal transsphenoidal surgery, *Clin. Neurol. Neurosurg.* 190 (2020), <https://doi.org/10.1016/j.jclineuro.2020.105743>.
- [150] C. Laule, I.M. Vavasour, S.H. Kolind, D.K.B. Li, T.L. Trabulsee, G.R.W. Moore, A. L. Mackay, Magnetic Resonance Imaging of Myelin, *Neurotherapeutics*. 4 (2007) 460–484, <https://doi.org/10.1016/j.nurt.2007.05.004>.
- [151] R.A. Horsch, J.C. Gore, M.D. Does, Origins of the ultrashort-T2 1H NMR signals in myelinated nerve: A direct measure of myelin content?, *Magn. Reson. Med.* 66 (2011) 24–31, <https://doi.org/10.1002/mrm.22980>.
- [152] M.J. Wilhelm, H.H. Ong, S.L. Wehrli, C. Li, P.-H. Tsai, D.B. Hackney, F.W. Wehrli, Direct magnetic resonance detection of myelin and prospects for quantitative imaging of myelin density, *Proc. Natl. Acad. Sci.* 109 (2012) 9605–9610, <https://doi.org/10.1073/pnas.1115107109>.
- [153] A. MacKay, K. Whittall, J. Adler, D. Li, D. Paty, D. Graeb, In vivo visualization of myelin water in brain by magnetic resonance, *Magn. Reson. Med.* 31 (1994) 673–677, <https://doi.org/10.1002/mrm.1910310614>.
- [154] H. Jang, M. Carl, Y. Ma, A.C. Searleman, S. Jerban, E.Y. Chang, J. Corey-Bloom, J. Du, Inversion recovery zero echo time (IR-ZTE) imaging for direct myelin detection in human brain: a feasibility study, *Quant. Imaging Med. Surg.* 10 (2020) 895–906, <https://doi.org/10.21037/qims.2020.04.13>.
- [155] S.-K. Lee, S. Bulumulla, F. Wiesinger, L. Sacolick, W. Sun, I. Hancu, Tissue Electric Property Mapping From Zero Echo-Time Magnetic Resonance Imaging, *IEEE Trans. Med. Imaging*. 34 (2015) 541–550, <https://doi.org/10.1109/TMI.2014.2361810>.
- [156] S.M. Lechner-Greite, F. Wiesinger, D.T.B. Yeo, M.R. Tarasek, Proton Density and T1 weighted Zero TE MR Thermometry, *US10,076,249 B2*, 2018.
- [157] E.M. Haacke, L.S. Petropoulos, E.W. Nilges, D.H. Wu, Extraction of conductivity and permittivity using magnetic resonance imaging, *Phys. Med. Biol.* 36 (1991) 723–734, <https://doi.org/10.1088/0031-9155/36/6/002>.
- [158] P. Soullier, A. Missoffe, K. Ambarki, J. Fellblinger, F. Odille, MR electrical properties imaging using a generalized image-based method, *Magn. Reson. Med.* 85 (2020) 762–776, <https://doi.org/10.1002/mrm.28458>.
- [159] D.I. Hoult, The principle of reciprocity in principle strength calculations - A mathematical guide, *Concepts Magn. Reson.* 12 (2000) 173–187, [https://doi.org/10.1002/1099-0534\(2000\)12:4<173::AID-CMR1>3.0.CO;2-Q](https://doi.org/10.1002/1099-0534(2000)12:4<173::AID-CMR1>3.0.CO;2-Q).
- [160] H.E. Cline, K. Hynnen, C.J. Hardy, R.D. Watkins, J.F. Schenck, F.A. Jolesz, MR Temperature Mapping of Focused Ultrasound Surgery, *Magn. Reson. Med.* 31 (1994) 628–636, <https://doi.org/10.1002/mrm.1910310608>.
- [161] M. Han, S.J. Scott, E. Ozhinsky, V.A. Salgaonkar, P.D. Jones, P.E.Z. Larson, C.J. Diederich, V. Rieke, R. Krug, Assessing temperature changes in cortical bone using variable flip-angle ultrashort echo-time MRI, in: *AIP Conf. Proc.*, 2017: pp. 060001. doi:10.1063/1.4977625.

- [162] S. Ogawa, T.M. Lee, A.R. Kay, D.W. Tank, Brain magnetic resonance imaging with contrast dependent on blood oxygenation, *Proc. Natl. Acad. Sci. U. S. A.* 87 (1990) 9868–9872, <https://doi.org/10.1073/pnas.87.24.9868>.
- [163] N.L. Damestani, D.J. Lythgoe, O. O'Daly, F. Wiesinger, A.B. Solana, S.C.R. Williams, F. Zelaya, Silent functional MRI for novel sound discrimination using the auditory oddball paradigm, in: *Proc. Intl. Soc. Mag. Reson. Med* 28, 2020: p. 3894, <https://archive.ismrm.org/2020/3894.html>.
- [164] S. Mangia, R. Chamberlain, F. De Martino, S. Moeller, C. Corum, T. Kim, C. Kalavagunta, S. Michaeli, M. Garwood, S.-G. Kim, K. Ugurbil, Functional MRI with SWIFT, *Proc. Intl. Soc. Mag. Reson. Med* 20 (2012) 326, <https://doi.org/10.13140/RC.2.1.4004.2728>.
- [165] L.J. Lehto, D. Idiyatullin, J. Zhang, L. Utecht, G. Adriany, M. Garwood, O. Gröhn, S. Michaeli, S. Mangia, MB-SWIFT functional MRI during deep brain stimulation in rats, *Neuroimage*. 159 (2017) 443–448, <https://doi.org/10.1016/j.neuroimage.2017.08.012>.
- [166] L.J. Lehto, P. Filip, H. Laakso, A. Sierra, J.P. Slopsema, M.D. Johnson, L.E. Eberly, W.C. Low, O. Gröhn, H. Tanila, S. Mangia, S. Michaeli, Tuning Neuromodulation Effects by Orientation Selective Deep Brain Stimulation in the Rat Medial Frontal Cortex, *Front. Neurosci.* 12 (2018) 899, <https://doi.org/10.3389/fnins.2018.00899>.
- [167] J. Paasonen, H. Laakso, T. Pirttimäki, P. Stenroos, R.A. Salo, E. Zhurakovskaya, L.J. Lehto, H. Tanila, M. Garwood, S. Michaeli, D. Idiyatullin, S. Mangia, O. Gröhn, Multi-band SWIFT enables quiet and artefact-free EEG-fMRI and awake fMRI studies in rat, *Neuroimage*. 206 (2020), <https://doi.org/10.1016/j.neuroimage.2019.116338>.
- [168] A.B. Solana, J.A. Hernández-Tamames, E. Manzanedo, R. García-Álvarez, F.O. Zelaya, F. del Pozo, Gradient induced artifacts in simultaneous EEG-fMRI: Effect of synchronization on spiral and EPI k-space trajectories, *Magn. Reson. Imaging*. 32 (2014) 684–692, <https://doi.org/10.1016/j.mri.2014.03.008>.
- [169] L.N. Tanenbaum, A.J. Tsiouris, A.N. Johnson, T.P. Naidich, M.C. DeLano, E.R. Melhem, P. Quarterman, S.X. Parameswaran, A. Shankaranarayanan, M. Goyen, A.S. Field, Synthetic MRI for clinical neuroimaging: Results of the magnetic resonance image compilation (MAGiC) prospective, multicenter, multireader trial, *Am. J. Neuroradiol.* 38 (2017) 1103–1110, <https://doi.org/10.3174/ajnr.A5227>.
- [170] G. Wang, E. Gong, S. Banerjee, D. Martin, E. Tong, J. Choi, H. Chen, M. Wintermark, J.M. Pauly, G. Zaharchuk, Synthesize High-Quality Multi-Contrast Magnetic Resonance Imaging From Multi-Echo Acquisition Using Multi-Task Deep Generative Model, *IEEE Trans. Med. Imaging*. 39 (2020) 3089–3099, <https://doi.org/10.1109/TMI.2020.2987026>.
- [171] I. Corcuera-Solano, A. Doshi, P.S. Pawha, D. Gui, A. Gaddipati, L.N. Tanenbaum, Quiet PROPELLER MRI techniques match the quality of conventional PROPELLER brain imaging techniques, *Am. J. Neuroradiol.* 36 (2015) 1124–1127, <https://doi.org/10.3174/ajnr.A4235>.
- [172] J. Rösch, A. Mennecke, M. Knott, A. Doerfler, D.M. Grodzki, Quiet FLAIR at 7T MRI, *Invest. Radiol.* 55 (2020) 722–726, <https://doi.org/10.1097/RLI.0000000000000694>.
- [173] J. Hutter, A.N. Price, L. Cordero-Grande, S. Malik, G. Ferrazzi, A. Gaspar, E.J. Hughes, D. Christiaens, L. McCabe, T. Schneider, M.A. Rutherford, J.V. Hajnal, Quiet echo planar imaging for functional and diffusion MRI, *Magn. Reson. Med.* 79 (2018) 1447–1459, <https://doi.org/10.1002/mrm.26810>.
- [174] J. Rösch, M. Ott, B. Heismann, A. Doerfler, T. Engelhorn, K. Sembritzki, D.M. Grodzki, Quiet diffusion-weighted head scanning: Initial clinical evaluation in ischemic stroke patients at 1.5T, *J. Magn. Reson. Imaging*. 44 (2016) 1238–1243, <https://doi.org/10.1002/jmri.25228>.
- [175] A. Nordell, M. Lundh, S. Horsch, B. Hallberg, U. Åden, B. Nordell, M. Blennow, The acoustic hood: A patient-independent device improving acoustic noise protection during neonatal magnetic resonance imaging, *Acta Paediatr. Int. J. Paediatr.* 98 (2009) 1278–1283, <https://doi.org/10.1111/j.1651-2227.2009.01339.x>.
- [176] P. Glover, J. Hykin, P. Gowland, J. Wright, I. Johnson, P. Mansfield, An assessment of the intrauterine sound intensity level during obstetric echo-planar magnetic resonance imaging, *Br. J. Radiol.* 68 (1995) 1090–1094, <https://doi.org/10.1259/0007-1285-68-814-1090>.
- [177] A.L. Chartier, M.J. Bouvier, D.R. McPherson, J.E. Stepenosky, D.A. Taysom, R.M. Marks, The safety of maternal and fetal MRI at 3 T, *Am. J. Roentgenol.* 213 (2019) 1170–1173, <https://doi.org/10.2214/AJR.19.21400>.
- [178] M.J. Reeves, M. Brandreth, E.H. Whitby, A.R. Hart, M.N.J. Paley, P.D. Griffiths, J. C. Stevens, Neonatal Cochlear Function: Measurement after Exposure to Acoustic Noise during in Utero MR Imaging, *Radiology*. 257 (2010) 802–809, <https://doi.org/10.1148/radiol.10092366>.
- [179] R.D. Kok, M.M. De Vries, A. Heerschap, P.P. van den Berg, Absence of harmful effects of magnetic resonance exposure at 1.5 T in utero during the third trimester of pregnancy: A follow-up study, *Magn. Reson. Imaging*. 22 (2004) 851–854, <https://doi.org/10.1016/j.mri.2004.01.047>.
- [180] B. Strizek, J.C. Jani, E. Mucyo, F. De Keyser, I. Pauwels, S. Ziane, A.-L. Mansbach, P. Deltenre, T. Cos, M.M. Cannie, Safety of MR imaging at 1.5 T in fetuses: A retrospective case-control study of birth weights and the effects of acoustic noise, *Radiology* 275 (2015) 530–537, <https://doi.org/10.1148/radiol.14141382>.

Glossary

3D: Three dimensional
 ASL: Arterial Spin Labelling
 ASPIR: Adiabatic Spectral Inversion Recovery
 B_0 : Static magnetic field
 B_1 : Radio-frequency magnetic field
 B_1^+ : Transmit radio-frequency magnetic field
 B_1^- : Receive radio-frequency magnetic field
 BLAST: Back-Projection Low Angle Shot
 BOLD: Blood-Oxygen Level Dependent
 BRAVO: Brain Volume Imaging
 BW: Bandwidth
 CT: Computed Tomography
 CTA: Computed Tomography Angiogram
 dBA: A-weighted dB
 DBS: Deep Brain Stimulation
 DW: Diffusion Weighted
 EEG: Electroencephalography
 EP: Electrical Property
 EPI: Echo Planar Imaging
 FA: Flip Angle
 FID: Free Induction Decay
 FOV: Field of View
 FLASH: Fast Low-Angle Shot
 GRE: Gradient Echo
 FLAIR: Fluid Attenuated Inversion Recovery
 fMRI: Functional MRI
 FOV: Field of view
 HYFI: Hybrid Filling
 ihMT: Inhomogeneous Magnetization Transfer
 IR: Inversion Recovery
 mBIR4: modified B1 Insensitive Rotation Adiabatic RF Pulse with 4 segments
 MP: Magnetization Preparation
 MPAGE: Magnetization Prepared Rapid Gradient Echo
 MP2RAGE: Magnetization Prepared 2 Rapid Gradient Echoes
 MR: Magnetic Resonance
 MRA: Magnetic Resonance Angiography
 MRI: Magnetic Resonance Imaging
 MT: Magnetization Transfer
 PD: Proton Density
 PETRA: Pointwise Encoding Time reduction with Radial Acquisition
 RF: Radio Frequency
 RUFIS: Rotating Ultra-Fast Imaging Sequence
 QSM: Quantitative Susceptibility Mapping
 RT: Radiation Therapy
 SNR: Signal to Noise Ratio
 SPGR: Spoiled Gradient Echo
 SWIFT: Sweep Imaging with Fourier Transform
 T_1 : Longitudinal relaxation time
 T_2 : Transverse relaxation time
 T_2^* : Apparent transverse relaxation time
 TE: Echo Time
 TOF: Time of flight
 TR: Repetition Time
 UTE: Ultra-short echo time
 VFA: Variable Flip Angle
 WASPI: Water and fat-Suppressed proton Projection MRI
 ZTE: Zero Echo Time